


Lovibond® Colour Measurement Tintometer® Group



Indice de Gardner 2



Code: 134200

Reference liquid standards are assigned a nominal value and Lovibond Tintometer always endeavours to match this value as closely as possible. On occasion, it may not be possible to match the requested value exactly. The value achieved and the expected performance tolerances will, of course, always be reported.

Indice de Gardner 2

Flacon de 500 ml, certifié conforme à ISO/IEC 17025, peut être conservé pendant 12 mois.

Traçabilité complète, conforme aux normes internationales reconnues

Tous les flacons appartiennent à la catégorie de produits non dangereux, conformément aux directives UE

Chaque flacon est fourni avec des certificats complets, y compris fiche technique de sécurité.