Benchmark Scientific Inc.
PO Box 709
Edison, NJ 08818

DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: Benchmark PlateFuge™

Device Model Number(s): C2000 (-E) and Private label versions of these devices.

Relevant EC Directives:
Low Voltage Directive 2014/35/EU
EMC Directive 2014/30/EU

Harmonized Standards:
EN61326-1:2013
EN61000-3-2:2014
EN61000-3-3:2013
EN61000-4-2:2009
EN61000-4-3:2006+A2:2010
EN61000-4-4:2012
EN61000-4-5:2014
EN61000-4-6:2014
EN61000-4-8:2010
EN61000-4-11:2004
EN61010-1:2010

Manufacturer's Name:
Benchmark Scientific, Inc.
PO Box 709
Edison, NJ 08818

Date: October 18, 2022

Phone: (908) 769-5555 Fax: (732) 313-7007 Email: info@BenchmarkScientific.com