VENT or Filter Screw Cap

The TPP Click



VENT position at 12 o'clock = gas exchange

- Click:
  VENT screw cap is in position for gas exchange.
- Visual control gas exchange:
  rectangle is at 12 o'clock.



VENT position at 3 o'clock = gas tight

- Turn 90° degree clockwise over the click:
  VENT screw cap closes flask gas tight.
- Visual control of the gas tight position:
  rectangle is at 3 o'clock.


DOMINIQUE DUTSCHER SAS

VENT or Filter Screw Cap for Tissue Culture Flasks

Type	Product-No.	Version	Dimensions Ø x h mm	Material	Qty/ Bag pcs	Qty/ Case pcs
	90825	VENT for T-25	24 x 16	PE	10	40
	90826	filter for T-25	24 x 16	PE	10	40
	90850	VENT for T-150 / 300	39 x 24	PE	10	40
	90856	filter for T-150 / 300	39 x 24	PE	10	40
	90875	VENT for T-75	33 x 21	PE	10	40
	90876	filter for T-75	33 x 21	PE	10	40







1. 100 % cell harvesting
2. Bilateral marking area and volume scale
3. Angled neck