

Accumax Lab Devices Private Ltd.

Regd. Office:

228/1/4, Dantali Industrial Estate, Gota-Vadsar Road, Dantali,

Tal.: Kalol, Dist.: Gandhinagar, Gujarat, India, Pin:382721

T: +91-99244 77224, +91-99244 77225, +91-99244 77336 | www.accumaximum.com

CIN NO.: U33110GJ2009PTC056957 | www.neuation.com



EC DECLARATION OF CONFORMITY

We certify that the following designated product

Brand: Neuation

Category: Dry Bath

Model: iTherm D150-1

Date: 13/March/2023

Is in fully conformity with the harmonized standard(s)


EN 61326-1, EN 55011, EN 61010-1, EN IEC 61010-2-010.

Complies to essential applicable requirements

Electromagnetic Compatibility (EMC) 2014/30/EU,
Low Voltage Directive (LVD) 2014/35/EU,
RoHS Directive 2011/65/EU with inclusion of (EU) 2015/863.

The declaration is the sole responsibility of the manufacturer / importer

Place of Issue : Unit 1 : SEZ



Harsh Khatri
Certification Engineer, R&D

Unit 1: SEZ
Accumax Lab Devices Private Ltd.
Plot No. 15, 16 & 32
GIDC Electronic Park SEZ,
Sector 26, Gandhinagar,
Gujarat, India - 382026.

Unit 2: EOU
Accumax Lab Devices Private Ltd.
228/1/4,
Dantali Industrial Estate,
Gota-Vadsar Road, Dantali,
Tal.: Kalol, Dist.: Gandhinagar,
Gujarat, India - 382721.

Unit 3: DTA
Accumax Lab Devices Private Ltd.
194/6/2,
Dantali Industrial Estate,
Gota-Vadsar Road, Dantali,
Tal.: Kalol, Dist.: Gandhinagar,
Gujarat, India - 382721.

Unit 4: DTA
Accumax Lab Devices Private Ltd.
Survey No 140 / 141, Moje - Medha,
Tal.: Kadi, Dist.: Mehsana,
Gujarat, India - 384445.