

1. Part Number: 8050H-1500S

2. Photograph:



3. Product Information:

| | |
|---|---|
| Description: | SteriWare LiquiThief - 1500mm - Sterile |
| Use: | Single use, disposable liquid sampler |
| Material of Construction: | High Density Polyethylene (HDPE) - virgin <ul style="list-style-type: none"> • Conforms to FDA CFR 177.1520 • Conforms to EU Regulations 10/2011 • Conforms to EC Regulations 1935/2004 • BSE/TSE free O-ring Seal – virgin Silicone Rubber <ul style="list-style-type: none"> • Conforms to FDA CFR 177.2600 |
| Method of Construction: | Ultrasonic welding |
| Nominal Length | |
| Fully closed: | 1495mm |
| Fully extended: | 2830mm |
| Diameter of Outer Tube: | 21mm |
| Diameter of hole in tip: | 1.5mm |
| Width of Handle: | 103mm |
| Nominal Weight of Sampler: | 240g |
| Assembly & Primary Packing Environment: | Class 100,000 Cleanroom |
| Individually Bagged? | Yes (heat sealed PE/PET bag) |
| Method of Sterilisation: | Gamma Irradiation (minimum dose 25kGy) |
| Quantity per Box: | 20 |
| Recommended Storage Conditions: | Dry and ambient temperature |
| Shelf Life: | 5 years from the date of manufacture |
| ATEX: | This device has not been ATEX tested |

4. Change Control:

Sampling Systems operates a policy of continual improvement. To be kept informed of all major changes to any product in the SteriWare range please email your contact details to change.control@sampling.com



Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill, Warwickshire, B46 1JH, UK.
Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com

5. Country of Origin:

Made and packed in the UK.

6. Repeated Use:

This is a single-use, disposable sampler. Repeated use will lead to failure.

7. Approval:

Signed:



Name: Kuldeep Flora

Position: Quality Manager

Date: 6-Oct-20

DOMINIQUE DUTSCHER SAS