

Specification Sheet Specification Number: 1360A-1001 Revision: Date Issued:

02 9-Mar-20

1. Part Number:

1360A-1001

2. Photo:



3. Product Information:

Description: Material of Construction: Method of Construction: Surface Finish: **Overall Length:** Outer Tube Diameter: Number of Sampling Ports: Position of Sampling Ports: AccuPod[®] Sample Thief 1000mm, 1 port 316L stainless steel Welds ground and polished Better than 0.8 microns Ra (tube exteriors) 1000mm 25mm 1



All polishing compounds used in the manufacture of this sampler are

This sampler should be thoroughly cleaned before being used for

Nominal Weight: **BSE/TSE** statement:

Recommended Storage Conditions: First Use:

Note:

4. Approval:

This sampler is designed for use with AccuPod[®] inserts. AccuPod[®] inserts must be ordered separately. Contact Sampling Systems for details.

the first time

of vegetable origin

Dry and ambient temperature

2.6g

Signed:

Kuldeep Flora Name: Position: Quality Manager Date: 9-Mar-20



Sampling Systems Ltd. 4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com

SteriWare is a registered Trademark belonging to Sampling Systems Ltd.