

PDS No. 1150xx/ 115101	PRODUCT DATA SHEET				Page 1 of 1
Revision 05	Tube, 5 ml, PS				 greiner bio-one
	Item-No. 1150xx / 115101				
Valid for Item-No.:	115001	115061	115070	115071	115101

1.	Description / Specification	
1.1	Description	PS tube 115001: Immuno Tube, MICROLON®200, medium binding, round bottom 115061: Immuno Tube, MICROLON®600, high binding, round bottom 115070: Immuno Tube, MICROLON®200, medium binding, round bottom with star 115071: Immuno Tube, MICROLON®600, high binding, round bottom with star 115101: without treatment, round bottom Appropriate grip stopper, Cat.-No. 303321
1.2	Dimensions	See Customer Drawing
1.3	Volume	Working volume: 4 ml
1.4	Material / Resin	PS (Polystyrene), free of heavy metal
1.5	Colour	Clear
1.6	Sterilization	No
1.7	Quality Control	<u>Raw Material-Control:</u> 115101: physical testing 115001, -061, -070, -071: physical and immunological testing <u>Product-Control:</u> testing of attributable and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	-
2.2	Temperature range	-20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	5800 x g: swinging-bucket rotor 5800 x g: fixed-angle rotor
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	115001, -061, -070, -071: 4 years after month of production 115101: n/a
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	250
3.2	Pieces / Box	2000
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 04	Date 17 April 2015	Date 20 April 2015	Date 20 April 2015	
Date 01.08.2014	Name S. Kaelberer	Name Dr. U. Honisch	Name A. Schulz	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.