Manual of LabV Series





Note:

Please read the manual carefully before operating the product.



Warning:

Connect the power cord to the wall socket directly, and avoid using the extended electric wire.

P

- If the power cord or plug had wear and other damage, please disconnect the plug. (Hold the plug instead of the wire)
- If following situations happened, please turn off the power supply and disconnect the plug. (Hold the plug instead of the wire)
 - 1. Fluid splash on the pump.
 - 2. You think the pump need to maintain or repair.
- > The user's power socket must have ground wire, and have reliable grounding.

Note: The foot pedal switch and other external control plugs must be connected or disconnected in the power-off status to prevent the external control interface from being burned.

Table of contents

1. Product Introduction	1 -
2. Product Appearance	1 -
3. Keyboard Instruction	2 -
4. Operation Interface Structure	2 -
5. Main Functions Operation Process	16 -
6. External Control Interface Instruction	19 -
7. Technical Specification	26 -
8. Functions and Features	27 -
9. Dimension Drawing	28 -
10. Maintenance	- 33 -
11. Warranty and After-sales Service	- 33 -



1. Product Introduction

LabV series product are the flow measurement type intelligent peristaltic pump, 4.3 inch color touch screen control; Animation shows working state; flow data, setting parameters and system settings are displayed in the same screen. Intelligent calibration and online micro adjusting function; three measurement modes: fixed volume measurement, dispensing, timer start and stop. It can load different pump heads, multiple external methods are optional. It is the ideal choice for laboratory, equipment supporting and industrial production.

Product mode includes: LabV1, LabV3, LabV6, LabV1-II, LabV3-II, LabV1-III, LabV3-III, LabV6-III.

Suitable pump head: EasyPump pump head, AMC pump head (AMC1-AMC12), YZ1515x, YZ2515x pump head, MC series pump head (MC1-MC12). KD pump head.

2. Product Appearance





3. Keyboard Instruction



- **Stop Button**: Press stop button, pump stops working. Forbidden buttons in the main interface can be used. Keep pressing the button and turn on the pump power supply, that will initialize the pump and all the parameters will be lost.
- **Full Button**: At stopping state or transferring state, press this button, the pump will run with the highest speed. This button can be used for washing tube or fast filling liquid.
- **CW/CCW Button**: Press this button once, the motor will change running direction once. When the fixed volume measurement or dispensing are on, this button does not work.
- Start Button: Press this button, the motor starts running. When fixed volume measurement or dispensing function are on, press this button, the pump will start work with the function. Under the operation of fixed volume measurement, clicking this button again will pause current operation.



LabV Series Operation Interface Instruction



4.1 Booting Interface

After turn on the system, enter the Boot Screen, click anywhere or wait for 2.5 seconds it will enter the English main operation interface automatically.

4.2 Main Interface

Main Interface Composition as below:



- A. Speed/Flow Rate Display: In the flow rate mode, displays the current flow rate, the motor speed is displayed at the C frame. In the speed mode, displays the current set up speed, flow rate is displayed at the C frame. Click A to amend the flow rate or speed. When dispensing function is turned on, A is forbidden, not allowed to amend the flow rate or speed.
- **B. Real-time Dynamic Display:** Real-time display of the current running status, dynamic display of the running results.
- C. Real-time Parameter Display: Display the current running state and setting up parameter. When the fixed volume measurement turns on, displays the fixed volume measurement parameter; when the dispensing function turn on, display the dispensing parameter. When these two functions turn off,

Innofluid

displaying parameter is all zero.

- **D.** Setting Parameter Display: Display the fixed volume measurement, dispensing state information, the model of pump head and tube size.
- E. Date and Time Display: Display the current date and time, you can change it in the system settings. When it displays an alarm clock on the right side, it means the timer start and stop function had turned on.
- F. System Settings Button: Click this button, set up other parameters.
- **G.** Flow Calibration Button: Click this button to enter the flow rate calibration interface.
- **H.** Date & Time Button: Click this button to enter set up current date and time interface.
- I. Fixed Volume Measurement Button: Click this button, enter the fixed volume measurement interface.
- J. Dispensing Button: Click this button, enter dispensing interface.
- K. Timer Start and Stop Button: Click this button, enter timer start and stop interface.



4.3 Numeric Keypad Input Interface

Numeric keypad input interface as below:



Input Information Display: The information displayed is the current operation object.

Input Data Display: Display the current input data, range is 0.01-9999.

Unit Display: Display input units when input flow rate or volume.

Input Digital Area: Numeric keypad.

Unit/Clr Button: When input flow rate or volume, this button is unit switch, you can choose different units. When it is Clr, you can clear the current input data.

Backspace Button: Delete an input digital.

ESC Button: Cancel the current input, back to previous interface.

ENT Button: Confirm the current input.



4.4 The Basic Configuration Interface

The basic configuration interface:



Click the **Pump Head** and **Tubing Size** to choose the pump head and tubing.

Reference flow rate displays the maximum and minimum flow rate with the current pump head and tubing.

Click the Flow rate mode or Speed mode button to choose the working mode. When you choose the flow rate mode, the flow rate is adjustable, the speed will change with the flow rate. When you choose the speed mode, the speed is adjustable, the flow rate will change with the rotating speed.

Click the confirm button when you had finished choosing parameter then back to the main interface.

Note that: When the pump come with two pump heads, the output of two pump heads are connected to one channel with Y type connector, then you will need to choose 2* pump head model; if the two pump heads use as two channels, then you need to choose single pump head model number.

For example, the pump come with two EasyPumpI, and output connect with Y type connector to one channel, then when choose pump head need to select 2*EasypumpI, as in below picture:



2*EasyPumpl/III

In other cases, such as: the pump come with one pump head EasypumpI, or with two EasypumpI use as two channels, or with 3 or 4 EasypumpI pump heads, need to select single pump head EasypumpI, as in below picture:

EasyPumpl/III

4.5 Back Suction Angle Interface

The back suction angle interface as below:



Click the **System Settings** button in the main interface, then click **Back Suction** button to enter the back suction angle setting interface. Click **angle** button, pop up the numeric keyboard for input the suction angle, enter back suction angle then click OK. If click the **Cancel** button, it will back to the system settings main interface.



4.6 External Setting Interface

External Setting Interface as below:



Click System Settings button in the main interface, then click External control to enter External Settings interface.

- a. There are two types of signals for external control motor start/stop and direction: Level mode and Pulse mode. Connection interface refer to the external control interface instruction.
- b. Various external control modes are independently set on switches, which will only work after the corresponding external control function is turned on.



4.7 External Speed Control Setting Interface

External Speed Control Setting Interface as below:



Click **System Settings** button in the main interface, then click **External Speed Control** button to enter External Speed Control Settings interface.

According with external input signal to set the analog signal: 0-5V, 0-10V or 4-20mA. Between analog signal voltage range and motor speed, there is linear relationship.(when the working speed limit is off).

After turn on the maximum working speed limit, the motor speed will be limited. For example, if 0V to 0rpm, 5V to 600rpm (then 2.5V should be 300rpm). But if the maximum working speed limit is 300rpm, when external input analog signal is 2.5V, the motor speed is 300rpm. After the input signal exceeds 2.5V, the motor speed remains unchanged at 300 rpm.



4.8 Communication Setting Interface



Click **System Settings** button in the main interface, then click **Communication** button to enter Communication Settings interface.

This pump support Modbus--RTU Mode. Please select baud rates and communication interface (RS485/RS232). Click **Slave No.** button to enter peristaltic pump address No.(range:1-32), select communication enable is **ON**. Then this pump can be communication with master, receiving master signal.

Note: After settings, the peristaltic pump only receives communication control when in the main interface, it is invalid for the communication control in the other settings interface.



4.9 Flow Rate Calibration Interface

Flow Rate Calibration Interface as below:



The top left corner displays the function, when fixed volume measurement turn on, displays fixed volume; When dispensing turned on, displays dispensing. Others displays transferring mode.

If dispensing turned on, the target volume and running time are set up parameter, unable to amend. Other modes, the running time is 60s, you can click the run time button to amend the running time.

Before the pump works, need to calibrate the flow rate to ensure the transferring or dispensing accuracy

Process as below:

- (1) Confirm the running time, if dispensing function, the running time is set up time, unable to change.
- (2) Click **Test** button to start the test, countdown displays the run time, it will stop automatically, and display numerical keyboard, input the actual volume, then it will ask whether continue test (suggest more than 3 times), choose **Yes**, the pump will test again, choose **No**, back to the calibration interface.
- (3) After click the Test button, during the pump running, you can click the Stop

Innofluid

button to stop the test.

- (4) After finish the tests, the actual volume area display the average data, click the CAL button, prompt the calibration is successful.
- (5) If request higher accuracy, you can click **Add** and **Dec** button to micro adjust the flow rate, to reach high accuracy transferring and dispensing.
- (6) Click **Reset** button, restore to the factory default calibration parameters.

Online Micro Adjust Volume Process:

Flow Rate Transferring Mode: If the actual flow rate during the production process is bigger or smaller than the set up flow rate, you can micro adjust the flow rate online without affecting the product line.

Fixed Time and Volume Mode: If the dispensing volume is bigger or smaller than the set up volume, you can micro adjust the volume online, no need to stop the pump.

Fixed Volume Measurement Mode: Do not support online micro adjust function.

- Click the Calibration button from the main interface, enter the flow rate calibration interface.
- > Now only the Add, Dec and Esc button are usable, other buttons are forbidden.
- > Click Add or Dec button to micro adjust the flow rate or volume.

S Innofluid

4.10 Date & Time Interface

Setting Date & Time Interface as below:



In that interface, you can set the current date and time, it will display at the top right corner. Click **Set Date** button, come out the **Set year** numeric keypad, the range of the year is **1970-2099**. After set up the year, then set the month and date.

Click **Set Time** button, come out the numeric keypad, set the hour, minute and second.



4.11 Fixed Volume Measurement Interface

Fixed Volume Measurement Interface as below:



After turn on this function, the peristaltic pump will measure the volume automatically, when set up volume finished, the pump will stop working automatically. The flow rate can be changed during the pump working.

Click the **Fixed Volume** button, set **ON** to turn on this function. Click **Set Volume**, to input volume, the unit can be mL or L, range is 0.01mL to 9999L. The prompting frame display the needed time to finish the volume with set up flow rate. The maximum time is 9999min, when more than 9999min, the system will warn.

4.12 Dispensing Interface

Dispensing Interface as below:



Innofluid

After turn on this function, the pump will enter dispensing mode.

Peristaltic pump transfer fixed volume in fixed time, transfer times are the Repeat times, click suspend time button then input suspend time, prompt box display current diameter, after click the OK button, click the start button, the pump begin dispensing according to the diameter. P.



4.13 Timer Start and Stop Interface

Start pump and stop time can be set in any time on this function. After the current time reaches to the setting time, it will automatically execute start/stop function or stop the motor.

When the Fixed volume measurement or fixed time and volume function is turned on, the timer stop function is unavailable.



5. Main Functions Operation Process

5.1 Flow Rate Transferring Function



Note: Please refer the flow rate calibration interface instruction for flow rate

calibration process.

5.2 Fixed Volume Measurement Function



Note: Please refer the flow rate calibration interface instruction for flow rate calibration process.



5.3 Dispensing



Note: Please refer the flow rate calibration interface instruction for flow rate

calibration process.

5.4 Timer Start And Stop Function



S Innofluid

0

Under the flow rate transferring mode, set the pump start at 8:30 a.m. from Monday to Friday, stop at 5:30 p.m., the process as below:



Click Start Time, set the start time is 8:30 a.m., turn the button to ON.

Click **Custom**, come out the repeat date window, as below:

🗹 Monday 🗹 Friday
🗹 Tuesday 🔲 Saturday
Wednesday 🗖 Sunday
Thursday OK



6. External Control Interface Instruction

External control interface as below:





Pin	Definition	Explanation	Note
1	0-5V	0V to 5V voltage signal input	
1 0-3 V		terminal	
2	0.10V	0V to 10V voltage signal input	Analog signal input
2 0-10V		terminal	terminal
3	4-20mA	4-20mA current signal input terminal	
4	I_/V_ Analog signal negative terminal		
5	1 1		/
	/		The passive switch or
	D/C1	External control start/stop signal with	foot pedal switch can be
6	N/51	passive signal input	connected with the
			terminal. Set the validity



6

			of this input in external
			setting interfacefoot
			pedal option.
7	D/S2	External start/stan signal input	Active signal 5-24VDC
/	N/52	External start/stop signal input	input
0	NC		Active signal 5-24VDC
0	NC	External full speed signal input	input
0	GW/GGW		Active signal 5-24VDC
9	Cw/CCw	External direction signal input	input
10	/	/	/
11	/	/	/
12	/	1	/
13	/	1	/
14	/	/	/
15	/		/
16	OUT1	Operating status output terminal 1	Open collector output
17	OUT2	Operating status output terminal 2	/
10	0014		External Power Supply
18	COM	Provides voltage for logic outputs	Positive
19	GD2	Connected to the Internal GD1	/
20	1	/	/
21		/	/
22	Ń	/	/
23	+5V	Internal 5V output positive terminal	
24	GD1	Internal 5V output negative terminal	
	0	Connected internally to GD1 and	Internal 5V output
25	0V	GD2	





External control wiring and function description

Analog: 0-5V
$\begin{array}{c} 13 & 4 & 1 \\ 0 & 0 & 0 & 0 & 0 & 0 & 0 \\ 25 & 14 \\ \hline Analog: 0-10V \\ 13 & 4 & 2 \\ 0 & 0 & 0 & 0 & 0 & 0 \\ \hline 13 & 4 & 2 \\ 0 & 0 & 0 & 0 & 0 & 0 \\ \hline 13 & 4 & 2 & 0 & 0 & 0 \\ \hline 13 & 4 & 2 & 0 & 0 & 0 & 0 \\ \hline 13 & 4 & 2 & 0 & 0 & 0 & 0 & 0 \\ \hline 14 & 14 & 14 & 14 & 14 & 14 \\ \hline 15 & 14 & 14 & 14 & 14 & 14 & 14 & 14 \\ \hline 16 & 14 & 14 & 14 & 14 & 14 & 14 & 14 & $
$\frac{1}{14}$





Innofluid





2 XS6 Connector

Innofluid





Pin	Pin Definition	Explanation	Note
1	R/S1		The passive switch or foot pedal
		External control	switch can be connected with
		start/stop signal with	the terminal. Set the validity of
		passive signal input	this input in external setting
			interfacefoot pedal option.
2	GD1	Internal 5V output	1
		negative terminal	
3 DB	9 Connector	5 1 00000 0000	2011

3 DB9 Connector

Innofluid



Pin	Pin Definition	Explanation	Note
1	1	1	/
2	RS232-TXD	Signal transmitted by	Choose RS232 in the
	G	peristaltic pump, received	Communication setting
		by upper computer	interface, this terminal is
3	RS232-RXD	Signal received by	active.
		peristaltic pump,	
		transmitted by upper	
		computer	
4	/	/	/
5	GND	Communication ground	
		port	
6	/	/	/



7	GND1	RS485 sig	nal ground		
8	DC495 D	Connect	RS485	B-	Choose RS485 in the
	К5485-В	terminal			communication setting
9	DC405 A	Connect	RS485	A+	interface, this terminal is
	K5465-A	terminal			active.
Communication wiring and function description					

Communication wiring and function description

Signal and wiring	Function description
Computer port RS232 communication interface GND GN	RS232 communication interface: Choose RS232 in the Communication setting interface, this terminal is active.
RS485 Communication interface 5 1 $\bigcirc \bigcirc $	RS485 Communication Interface: Choose RS485 in the Communication setting interface, this terminal is active.

Note: Whether to choose RS232 or RS485, the communication protocol is standard



MODBUS protocol.

7. Technical Specification

7. Technical	Specification		C
Flow rate resolution	0.01ml/min	Power supply	AC 220V±10% 50Hz/60Hz(standard) AC 110V±10% 50Hz/60Hz(optional)
Operation mode	Touch screen and mechanical keypad	External control	Switch signal
External speed control	0-5V, 0-10V, 4-20mA for option	External control	Passive switching signal: Foot pedal switch Active switching signal: 5-24V universal
Communication	RS232/RS485, support Modbus (RTU mode)	Output port	Output motor running status(Open collector output)
Back suction angle	0-360°	Protection IP31 rate	
	LabV1, LabV1-II, LabV1-III		0.1-150rpm
Speed range	LabV3, LabV3-II, LabV3-III	0.1-350rpm	
	LabV6, LabV6-III	0.1-600rpm	
Power	LabV1, LabV1-II, LabV3, LabV3-II, LabV6	<50W <80W	
consumption	LabV1-III, LabV3-III, LabV6-III		



Motor type	LabV1, LabV1-II, LabV3, LabV3-II, LabV6		Stepper motor
	LabV1-III, LabV3-III,		Closed loop stepping motor
	LabV6-	III	Closed loop stepping motor
Temperature	0-40°C	Humidity	<80%

8. Functions and Features

- 4.3 inch color touch screen control, animation shows working state, the flow volume and motor speed are displayed in the same screen.
- Intelligent calibration function, it can calibrate the flow rate and dispensing volume, ensure the flow accuracy, suitable for high accuracy transferring liquid.
- On-line micro adjusting function, it can adjust the flow rate during production progress, to avoid the filling errors because of tubing fatigue and elasticity decreased.
- Accurate angle control technology, reach high precision dispensing and measurement.
- Fixed volume measurement function: After turn on this function, the peristaltic pump will measure the liquid volume automatically, it will stop automatically after the volume reaches the set value. During this process, the flow rate can be changed. It is suitable for liquid metering in the laboratory or quantitative feeding in the chemical reaction process, etc.
 - Fixed time and volume function: After turn on this function, the peristaltic pump will transfer fixed volume within set time. It is suitable for liquid dispensing in laboratory and industrial production.
- Timer start and stop function: Pump starts and stops time can be set any time to realize automation control.
- > Power down memory function, store the running parameters in time, safe and

Innofluid

reliable.

- > Fast fluid-filled function, can wash the tubing and also fill fluid in the tubing.
- High torque and low power loss, it can load several pump heads or multichannel pump head, meet different application requests.

> External control start and stop, convenient for equipment supporting.

9. Dimension Drawing

Unit: (mm)

9.1 Single pump head



YZ1515x pump head







EasyPump pump head





KD Pump Head

9.2 LabV series product

~~~



LabV+YZ1515X/2515X pump head

Note: For each additional pump head in series, the longitudinal dimension shall be increased by 55mm.





LabV + MC5 pump head

Note: For each additional channel, the longitudinal dimension shall be increased by 10mm.



LabV + AMC2 pump head

Note:For each additional channel, the longitudinal dimension shall be increased by 10mm.





LabV+EasyPump pump head

Note:For each additional pump head in series, the longitudinal dimension shall be increased by 61mm.



#### LabV+KD Pump head

Note:For each additional pump head in series, the longitudinal dimension shall be increased by 63mm.

### **S** Innofluid

#### 10. Maintenance

- Check the running status of machine before starting it, normal operation can be put into use.
- > Check for leakage, and correct fault which can be appeared.
- > Clean liquid overflowed from the pump in time.
- Please turn off the power supply and unplug the power socket (Hold the socket instead of power cord) when liquid splashed on pump. Check whether liquid flows into the machine, if it does, please contact the manufacture.
- The foot pedal switch and other external control plugs must be connected or disconnected in the power-off status to prevent the external control interface from being burned.
- > The user's power socket must have ground wire, and have reliable grounding.
- This product has no waterproof measures. Please take protective measures when using in water environment.
- This product does not have special certification such as medical certification. When it needs to be used in special fields such as medical and military, please self-certify.
- If the pump does not use for a long time, please clean it and keep it in dry and ventilated environment.
- The company shall not bear the direct and indirect losses caused by the malfunction or improper operation of this product.

#### 11. Warranty and After-sales Service

We support 3 years warranty for the pumps, subject to the exceptions below. Our company shall not be liable for any loss, damage, or expense directly or indirectly related to or arising out of the improper use of its products.

This warranty does not obligate our company to bear any costs of removal, installation, transportation, or other charges which may arise in connection with a warranty claim.

## **S** Innofluid

If the pump fails during the warranty period, after confirmation by our technical department, we will provide spare parts free of charge. Customers will need to bear the shipping cost.

#### **Exceptions:**

- The warranty shall not apply to repairs or service necessitated by normal wear and tear or for lack of reasonable and proper maintenance.
- > All tubing and pumping accessories as consumable items are excluded.
- Electrical surge as a cause of failure is excluded.
- > Chemical attack is excluded.

> Improper operation or man-made damage as a cause of failure is excluded.

pommula lie purscutte she

pommula lie purscutte she

Innofluid Co., Ltd.

Add: Building 10, No. 860, Xinyang Road, Lingang New Area, Pilot Free Trade Zone, Shanghai, China.

Manufacturer: Baoding Shenchen Precision Pump Co., Ltd.

Ç

Tel: 0086-312-5958380

Fax: 0086-312-6780636

Website: www.innofluid.com

Email: info@innofluid.com

MADE IN CHINA