

# Declaration of Conformity

|                                     |  |
|-------------------------------------|--|
| <b>Manufacturer:-</b>               | GRANT INSTRUMENTS (CAMBRIDGE) LTD,<br>Shepreth,<br>Cambridgeshire<br>SG8 6GB |
| <b>Equipment Name/Type Number:-</b> | XUBA1, XUBA3, XUB5, XUB10, XUB10, XUB12, XUB18, XUB25                        |
| <b>Description of Equipment:-</b>   | Analog and digitally controlled ultrasonic cleaning baths                    |
| <b>Directives:-</b>                 | EMC Directive 2004/108/EC<br>Low Voltage Directive 2006/95/EC                |
| <b>Including Accessories:-</b>      | Supplied mains cordset, Basket, lid, drain pipe                              |
| <b>CE marking first applied:-</b>   | 2010   |

|                               |  |
|-------------------------------|--|
| <b>Applied Standards:-</b>    |  |
| <b>Harmonized Standards:-</b> | BS EN 61326-1:2006 Electrical Equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements<br><br>BS EN61010 Part 1:2001 Safety requirements for electrical equipment for measurement, control and laboratory use |

*The products comply with the above directives when used with the supplied mains cord-set. Compliance cannot be assumed if an alternative mains cord-set is used.*

I declare that this apparatus conforms to the requirements of the above Directive(s)



.....  
Ludo Chapman  
Managing Director  
Grant Instruments (Cambridge) Ltd.

Dated.....19/12/12.....