## **EU Declaration of Conformity**

We the Manufacturer GRANT INSTRUMENTS (CAMBRIDGE) LTD,

Shepreth,

Cambridgeshire,

SG8 6GB

## Declare under our sole responsibility that the following Product(s);

Description of equipment Refrigerated and heating bath circulator

Model(s) ECOCOOL 150R, ECOCOOL 150R AUS, T ECOCOOL

150R

Serial no.(s) WWXXYYZZZ where

WW= product reference: 09, 0D

XX= Year of manufacture. e.g 16 =2016

YY= Week number of manufacture from 1 to 52. ZZZ= Number of unit made during the week number

above. From 1 to 999

Including Accessories cordset, mains cable link, serial link cable

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation;

Directive(s) EMC Directive 2014/30/EU

ROHS Directive 2011/65/EU (including 2015/863)

Machinery Directive 2006/42/EC

## Conformity was checked in accordance with the following harmonized standards;

EMC BS EN 61326-1:2013

Safety BS EN 61010-1:2010+A1:2019

BS EN 61010-2-010:2020 BS EN 61010-2-011:2019

BS EN 63000:2018

Kevin Belgrove Dated: 31-03-2022

**Technical Director** 

RoHS

Grant Instruments (Cambridge) Ltd. CE mark first applied: 31-03-2022

NOTE: Product complies with the listed directives only when used with the supplied accessories.

This device complies with Part 15 of the FCC rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference and (2) this device must accept any interference received, including interference that may cause undesired operation.