

Benchmark Scientific Inc.
PO Box 709
Edison, NJ 08818

DECLARATION OF CONFORMITY

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

Device Name: Accuris™ MyView™ Compact UV Transilluminator

Device Model Number(s): E3100, E3100-E and Private label versions of these devices.

Relevant EC Directives and Standards:

LOW VOLTAGE DIRECTIVE 2014/35/EU EN61010-1: 2010

EMC DIRECTIVE 2014/30/EU EN61326-1: 2013

ROHS DIRECTIVE 2011/65/EU EN50581: 2012

Manufacturer's Name: Benchmark Scientific, Inc.
PO Box 709
Edison, NJ 08818

Date: July 20th, 2020



President