


PDS No. 18726x	PRODUCT DATA SHEET		Page 1 of 1
Revision 05	Tube, 14 ml, Round Bottom, Two-Position Vent Stopper		
	Item-No. 18726x		
Valid for Item-No.:	187261 (sterile)	187262 (sterile)	

1.	Description / Specification	
1.1	Description	PP tube, 14 ml, round bottom, with two-position vent stopper, graduation and writing area, sterile
1.2	Dimensions	See Customer Drawing
1.3	Volume	Working volume: 12 ml
1.4	Material / Resin	Tube: PP (Polypropylene), free of heavy metal Stopper: LDPE (Low Density Polyethylene), free of heavy metal
1.5	Colour	Tube: translucent Stopper: translucent
1.6	Sterilization	SAL 10 ⁻³
1.7	Quality Control	- Raw Material-Control: physical testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens.
2.2	Temperature range	Tube: -196°C to +121°C Stopper: -50°C to +80°C
2.3	Autoclavability	Yes (without vent stopper)
2.4	Centrifugation, max. RCF	4800 x g: swinging-bucket rotor 34000 x g: fixed-angle rotor
2.5	Chemical Resistance	See homepage: https://www.gbo.com/en_INT/know-how-services/download-center.html
2.6	Shelf life	187261, -262: 5 years after month of production
2.7	Other Information	-

3.	Packaging	187261	187262
3.1	Pieces / Bag	1	25
3.2	Pieces / Box	800	1000
3.3	Lot-No.	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)	E YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	Certificate of Quality	Certificate of Quality

4.	Other Information
	-

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 04	Date 17 April 2015	Date 20 April 2015	Date 20 April 2015	
Date 01.08.2014	Name S. Kaelberer	Name Dr. T. Schreiber	Name A. Schulz	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.