

# TECHNICAL DATA SHEET

# EFL-1012

### **PRODUCT CODE**

EFL-1012

#### **PACKAGING**

Case of 1,000 (4 boxes of 250 bags)

### SHELF LIFE

5 years sterility guaranteed

# PRODUCT DESCRIPTION

Sterile sampling bag made of FDA approved virgin polyethylene tubing with tear off and easy pull tabs to open the bag. Bags are easily closed thanks to one round wire & one flat wire covered with white strips. Product is folded for convenient storage.

PRODUCT SPECIFICATI	ONS
Closure type	1 round & 1 flat wire with Safety Tabs
Closure material	Steel wires covered with FDA approved pressure sensitive tape with a backing of polypropylene or PVC.
Printed	Clear – Non-printed
MSDS	Not applicable for this product
Sterilisation	Sterility is assured due to extrusion temperatures exceeding 220°C which eliminates any present bacteria. During manufacturing all bags are sealed at both top and bottom in a continuous format before being cut into individual bags. This method assures that the bag's interior will never come in contact with its surroundings. Periodic validation tests, based on USP 40 – Sterility testing, are performed in our facility to guaranty the sterility of our processes.
RNase/DNase/Pyrogens	Products are free of RNase, DNase and pyrogens
Volume	76 oz (2250 mL)
Dimensions	10 in X 12 in (254 mm x 305 mm)
Thickness of material	4 mil( 102 micron)
FDA regulation	All TWIRL'EM sterile sampling bags comply to the applicable regulations of the Food and Drug Administration (21 CFR 177.1520 (c)3.1a, (c)3.2a and 178.2010.
EU regulation	All TWIRL'EM sterile sampling bags comply where applicable to EU No 10/2011 regulation for use in contact with food, subject to certain types of migration limitations and labeling requirements.
CFIA	All TWIRL'EM sterile sampling bags are also a solution which could be used in the CFIA Prevention Control Plan inspired by HACCP principles, a Canadian federal initiative from the Safe Food for Canadians Regulations (SFCR).

T: 450-649-7343 F: 450-649-3113 www.labplas.com service@labplas.com Updated: 07/10/2020