

**Bellco Glass, Inc.**  
**Specification Sheet**  
**Product #: 1985-00050**

**Description: Cell Screen, 50 Mesh**

- **Replacement Screen for Collector™ Tissue Sieve Homogenizer**
- **Screen Diameter: 1.812" ± 0.1"**
- **Size: 50 Mesh / 280 μ (microns)**
- **Wire Diameter" 0.009"**
- **Width Opening: 0.011"**
- **Open Area Percentage: 30.3%**
- **Tolerance Opening: ±0.0011"**

**The information contained in this specification sheet we believe is correct to the best of our knowledge. The recommendations and suggestions herein are made without guarantee or representation as to results. We recommend that adequate tests be made in your laboratory or plant to determine if this product meets all of your requirements.**



**BELLCO GLASS, INC.**



Bellco Glass, Inc. · 340 Edrudo Road · Vineland, NJ · 08360

Ver: 7/05