Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Product Description

Catalog Number: 6945-13L

Product Description: PYREX® Jar, Cylindrical, Chromatography, 8-3/4" OD x 8" High

Component Materials:
Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:
Capacity - 13.2 L Approximate
Product Dimensions - 222 mm OD x 457 mm High

Autoclaving:
This product will withstand repeated standard autoclave cycles of 121°C at 15 psi for 15 minutes. Due to the thick walls on this product it is CRITICAL that the heating and cooling cycles be slow and even.

Heating / Cooling:
This product MUST NOT be heated directly on a hot plate or over an open flame.

This product MUST NOT be cooled using liquid coolants.

Vacuum / Pressure:
This product is not designed for and is not rated for vacuum or pressure applications.

Sterilization:
Shipped Not Sterile

TSE/BSE:
Materials of animal origin are not used in the manufacture of this product.

Lot Number Designation:
Each production lot is uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Revision Date:
12 AUG 2013

Rev No: 1