

ATTESTATION OF CONFORMITY

Name and Address of Attestation Holder:

HANGZHOU ALLSHENG INSTRUMENTS CO., LTD.
BUILDING 9 NO.7 OF ZHUANTANG SCIENCE AND
TECHNOLOGY ECONOMIC ZONE, XIHU DISTRICT,
HANGZHOU CITY, ZHEJIANG 310024, CHINA.

Document Number

CE-1437-12-201123

Date of Issue:

20.11.2023

Name and Address of Manufacturer:

HANGZHOU ALLSHENG INSTRUMENTS CO., LTD.
BUILDING 9 NO.7 OF ZHUANTANG SCIENCE AND
TECHNOLOGY ECONOMIC ZONE, XIHU DISTRICT,
HANGZHOU CITY, ZHEJIANG 310024, CHINA.

Expiration Date:

20.11.2028

Test Report Number:

ALLSHENG-2023A12

Brand:

N/A

Test Required:

BS EN IEC 61010-1:2010+A1:2019
BS EN IEC 61010-2-010:2020
BS EN IEC 61010-2-101:2017
BS EN IEC 61010-2-081:2020
BS EN IEC 61326-1:2021
BS EN IEC 61326-2-6:2021
BS EN IEC 55011:2016+A11:2020

Product Name:

Nucleic Acid Purification System

Product Model:

Auto-Pure S32
MagnetaPure 32 Plus

The product meets the technical requirement of the above standards as mentioned in the reference test reports and hence fulfils the technical requirements of the following directives

2014/35/EU Low Voltage Directive **2014/30/EU Electromagnetic Compatibility Directive**

This document is only valid for the equipment and configuration described, in conjunction with the test data detailed above reference test reports. Document was issued on voluntary basis and does not imply meeting Notified Body conformity assessment procedure for the product.

The CE Mark, under the responsibility of the manufacturer or the importer, after completion of an EC Declaration of Conformity and compliance with all relevant EC Directives.

SIGNATURE