Arvensis Bioscience s.r.o.
Horny Bar 331
93033
Slovakia

E-mail: info@arvensis.uk

Certificate of Conformance

Issue Date	30.9.2026
Product Code	7014
Lot Number	7014280923
Expiry Date	Sept 2028

Arvensis certifies the product lot number detailed above has passed the following Quality Assessments:

Specification / Test	Result
Material Reaction tubes are 100% virgin medical grade polypropylene, inert, biocompatible, low binding.	PASS
Dimensional Check The above specified product conforms to all dimensional specifications.	PASS
Visual Inspection for Defects The above specified product is inspected under magnification for manufacturing defects.	PASS
Well Integrity Test Tube walls are free from holes. *Test method: electronic*	PASS
PCR Inhibitors The above specified product is free from any detectable PCR inhibitors. *Test method: quantitative PCR; Test sensitivity 2 copies of target gene*	PASS
RNase Test The above specified product is free from any detectable RNase contamination. *Test method fluorometric detection assay, test sensitivity 3.5e-7 Units/µl (0.5pg)*	PASS
DNase Test The above specified product is free from any detectable DNase contamination. *Test method: fluorometric detection assay, test sensitivity 5.0e-5 Units/µl (10pg)*	PASS
Human gDNA Test, Plant gDNA Test The above specified product is free from any detectable human and plant genomic DNA. *Test method: quantitative PCR; test sensitivity 2 copies of each target gene*	PASS
Heavy Metals Statement Our products do not contain any heavy metals.	PASS
Animal Free Statement (TSE/BSE) Our products do not contain or come into contact with any animal material.	PASS

Signed: _____

Certified by Peter Collins, Director