

CERTIFICATE

of Conformity EC Council Directive (EU) 2014/30 Electromagnetic Compatibility

Registration No.: AE 50636826 0001
Report No.: CN24R1CN 001
Holder: DLAB SCIENTIFIC CO.,LTD.
Yu An Road 31,
Airport Economic core zone,
Shun yi District,
Beijing, 101318
P.R. China
Product: Laboratory Equipment
(Centrifuge)


Type designation and serial number are listed on the next page(s)

This certificate of conformity is based on an evaluation of a sample of the above mentioned product. Technical Report and documentation are at the License Holder`s disposal. This is to certify that the tested sample is in conformity with all provisions of Annex I of Council Directive (EU) 2014/30. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity. The holder of the certificate is authorized to use this certificate in connection with the EC declaration of conformity according to the a.m. Directive.

Certification Body

Date: 2024-07-11

Gang Wang



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

The CE marking may be used if all relevant and effective EC Directives/Regulations are complied with.

CERTIFICATE

of Conformity EC Council Directive (EU) 2014/30 Electromagnetic Compatibility

Registration No.: AE 50636826 0001

Product: Laboratory Equipment
(Centrifuge)

Tested according to: EN IEC 61326-1:2021

Identification: Type Designation
D1008, Ezee, 9051000012, 901100X012
(X=0-9, X Stand for different client)
Serial No.:
A003726256-001



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

The CE marking may be used if all relevant and effective EC Directives/Regulations are complied with.

C E R T I F I C A T E



of Conformity
Low Voltage Directive 2014/35/EU

Registration No.: AN 50538960 0001

Report No.: CN21PN10 001

Holder: **DLAB SCIENTIFIC CO.,LTD.**
Yu An Road 31,
Airport Economic core zone,
Shun yi District,
Beijing, 101318
P.R. China

Product: Laboratory Equipment
(Centrifuge)

Identification: Type Designation: D1008 Ezee 9051000012 901100X012
(X=0-9,X stands for different client)
(DLAB)
Serial No.: Engineering sample
Remark: Refer to test report CN21PN10 001 for details.

This certificate of conformity is based on an evaluation of a sample of the above mentioned product. Technical Report and documentation are at the Licence Holder's disposal. This is to certify that the tested sample is in conformity with Annex I of Council Directive 2014/35/EU, referred to as the Low Voltage Directive. This certificate does not imply assessment of the series-production of the product and does not permit the use of a TÜV Rheinland mark of conformity. The holder of the certificate is authorized to use this certificate in connection with the EC declaration of conformity according to Annex IV of the Directive.

Certification Body



Date 14.04.2022

A handwritten signature in black ink, appearing to read 'Mai Miao'.

Mai Miao

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg