thermoscientific

G & M Procter Ltd.
Certificate of Analysis

PRODUCT	**PO1110A** **BRILLIANCE (TM) UTI CLARITY (TM) AGAR** **1X10PLT (9CM)**
LOT NUMBER	4574058
EXPIRY DATE	2026.11.26
PACKING DATE	2026.10.14
TEST DATE	2026.10.14
REPORTING DATE	2026.10.18

All testing in accordance with internally derived specifications, unless otherwise stated.

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Straw 2	Straw 2	SOP 178 Appearance and colour
pH (25°C)	7.1	6.8 - 7.2	SOP 53 pH
Fill Weight 19 ± 2 Grams	Conforms	Conforms	SOP 74 Fill volume weight check
Contamination @ 32 ± 2°C for >= 72 hours	Conforms	Within acceptable limits	SOP 167 Contamination Check at 22°C & 32°C

MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 50%-150% of the control medium and show the colonial appearance stated in the specification.

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Proteus mirabilis NCTC10975	92	84	Straw/brown cols	Straw/brown cols	SOP 151 Fertility of Specified Target Organism(s)(Agar)
Staphylococcus aureus ATCC®25923	77	70	Cream colonies	Cream colonies	SOP 151 Fertility of Specified Target Organism(s)(Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.

ilac-MRA UKAS TESTING 2727

Ian Snowball
Quality Manager
G & M Procter Ltd.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road, Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727

DOMINIQUE DUTSCHER SAS

thermoscientific

G & M Procter Ltd.
Certificate of Analysis

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Enterococcus faecalis ATCC®29212	89	92	Blue/green cols	Blue/green cols	SOP 151 Fertility of Specified Target Organism(s)(Agar)
Escherichia coli ATCC®25922	92	96	Pink cols	Pink cols	SOP 151 Fertility of Specified Target Organism(s)(Agar)
Enterobacter aerogenes ATCC®13048	89	91	Blue/purple cols	Blue/purple cols	SOP 151 Fertility of Specified Target Organism(s)(Agar)

DOMINIQUE DUTSCHER GAS

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.

ilac-MRA UKAS TESTING 2727

Ian Snowball
Quality Manager
G & M Procter Ltd.

thermoscientific

CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed
2026.10.18
Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

Ian Snowball
Quality Manager, G&M Procter Ltd

DOMINIQUE DUTSCHER SAS