OXOID Deutschland GmbH
Am Lippeglacis 4-8
D-46483 Wesel

CERTIFICATE OF ANALYSIS

PRODUCT **PO5040A** **CAMPYLOBACTER SKIRROW SELECTIVE MEDIUM**

LOT NUMBER 6458958

EXPIRY DATE 2026.12.15
TEST DATE 2026.09.08
REPORTING DATE 2026.09.11

General Characteristics	Results	Specification
Colour	Conforms	Ruby red
Appearance	Conforms	Transparent
pH	7.4	7.1 -7.5
Packaging / Presentation	Conforms	Conforms to specification
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Campylobact. jejuni ssp.jejuniATCC®33291	74	64	2-5mm, flat, grey, colonies
Escherichia coli ATCC®25922	1.6E+04	0	Complete inhibition (<= 10 cfu)
Staphylococcus epidermidis ATCC®12228	1.6E+04	0	Complete inhibition (<= 10 cfu)
Campylobacter coli ATCC®43478	<10000	Conforms	Good growth of characteristic cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

DOMINIQUE DUTSCHER SAS

The quality control methods meet requirements of ISO 11133.

The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAkkS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.