

# *Declaration of Conformity*



**Manufacture's Name**  
**Manufacture's Address**

Boekel Scientific  
855 Pennsylvania Blvd.  
Feasterville, PA 19053  
USA

**Testing Facility Name**  
**Testing Address**

Waltek Services Co., Ltd.  
8C, West Tower, Aidi Building  
No. 5004 Binhe Rd., Futian District  
Shenzhen, China

**Test Report Number(s)**

WT08082090-S-E-E

**Declares that the products listed conform to the requirements of the following European Directives:**

Name: Microplate Shaker  
Model(s): 270300 family  
FDA Product Classification: Class I  
FDA Product Code: JRQ

**EMC Directive 2004/108/EC**

**Conforming to the following standards:**

**EN 61010-1 (Product Safety)**

Certified compliance by Boekel Scientific, Inc.

**EN 61000-6-3:2007**

Certified by Waltek Services Co., Ltd.

**EN 61326-1:2006**

Certified by Waltek Services Co., Ltd.

**EN 61000-3-2:2006**

Certified by Waltek Services Co., Ltd.

**EN 61000-3-3:1995**

**+A2:2005**

**Waltek Services Co., Ltd.**  
8C, West Tower, Aidi Building  
No. 5004 Binhe Rd., Futian District  
Shenzhen, China

**Amy-Jo Frunzi, VP- Operations**  
**Feasterville, Pennsylvania USA**  
**January 2011**