

Technical Data Sheet

Doc No – Ab/TDS/GLP-147

Rev No : 04 Date 11/2019

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Measuring Cylinder – PP – Class B

Valid for Catalog No. P50301, P50302, P50303, P50304, P50305, P50306, P50307, P50308

1. Description/Specification

1.1 Description:

PP Measuring Cylinder

1.2 Design:

1.2.1 Body : Cylindrical with generous Pour spout for smooth of liquid While dispensing

1.2.2 Base : Hexagonal Base for stability

1.3 Performance Specifications:

Code	Capacity (ml)	Sub Division(ml)	Error limit (\pm ml)	Height (mm)
P50301	10	0.2	0.2	138
P50302	25	0.5	0.5	170
P50303	50	1	1	198
P50304	100	1	1	257
P50305	250	2	2	314
P50306	500	5	5	344
P50307	1000	10	10	412
P50308	2000	20	20	482

1.4 Material:

1.4.1 PP- Polypropylene (PP)

1.4.2 Comply with requirement of Directive 2002/72/EC and requirement of FDA Regulation 21 CFR 177.1520. olefins polymers.

1.4.3 Complies with the requirement of Indian Standard IS: 10910 on “Specification for Polypropylene and its copolymer for its safe use in contact with food stuffs, Pharmaceutical and Drinking water”

ABDOS

Shiv Ganga Industrial Estate,
Plot no.1 Lakeshwari, Bhagwanpur,
Roorkee – 247667 Uttarakhand, India
An ISO 9001 & 13485 Certified
A Division of Abdos

Regd. Office:
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903
GST : 05AAGCA8605A1Z7

P : +91 (0) 1332 235277
E : labtech@abdosindia.com
W: www.abdoslifesciences.com

Technical Data Sheet

1.4.4 Free from heavy metals according to ELV directive 2000/53/EC and its following amendments.

1.5 Color:

Translucent (Natural)

1.6 Graduations:

Molded in Large and easy to read graduation

Meet accuracy requirements of class B under “DIN 12681/ISO6706”

1.7 Sterilization:

Can be autoclaved at 15 Psi, 121 °C for 20 minutes

1.8 Quality control:

Testing of various characteristic in accordance with the valid Specification

1.9 other Information:

Reusable, can be autoclaved up to 10 autoclave cycle.

Pack Size: 3, 6 & 12 pieces/case

2. Basic Features

2.1 Made up of High purity Virgin Polypropylene (PP)

2.1 Temperature range: -30⁰C to +121⁰C

2.3 Autoclavability: Yes, Autoclavable at 15 psi and 121⁰C for 20 minutes

2.4 Chemical resistance: See website.

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121⁰C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50⁰C before removing product.


Quality Assurance Incharge
Abdos Labtech Pvt. Ltd.




ABDOS

Shiv Ganga Industrial Estate,
Plot no.1 Lakeshwari, Bhagwanpur,
Roorkee – 247667 Uttarakhand, India
An ISO 9001 & 13485 Certified
A Division of Abdos

Regd. Office:
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903
GST : 05AAGCA8605A1Z7

P : +91 (0) 1332 235277
E : labtech@abdosindia.com
W: www.abdoslifesciences.com