

# Technical Data Sheet

Doc No – Ab/TDS/CT-08  
Rev No : 08 Date 10/2021-22

## TECHNICAL DATA SHEET

**Abdos Labtech Pvt. Ltd.**

**MaxiRCF Trace Metal Free Centrifuge Tubes, PP/HDPE**

**Valid For Cat. No. P10480, P10481, P10482, P10483, P10484, P10485, P10486, P10487, P10488**

### **1. Description/Specification**

#### **1.1 Description:**

High Purity Virgin Polypropylene (PP) centrifuge tube, DNase, RNase, Human DNA, ATP, Pyrogen Free, Trace Heavy Metals Free, IVD & CE certified

#### **1.2 Design:**

1.2.1 Tube: Cylindrical Upper and conical Bottom/Self Stand

1.2.2 Tube Cap: Separate Threaded HDPE Cap

#### **1.3 Performance Specifications:**

1.3.1

<b>Code</b>	<b>Capacity</b>	<b>Outer Diameter(mm)</b>	<b>Height (mm)</b>	<b>No./Case</b>
<b>P10480</b>	15ml Bulk	16.5	121	500
<b>P10481</b>	15ml	16.5	121	500
<b>P10482</b>	15ml Sterile	16.5	121	500
<b>P10483</b>	50ml Bulk	29.2	116	500
<b>P10484</b>	50ml	29.2	116	500
<b>P10485</b>	50ml Sterile	29.2	116	500
<b>P10486</b>	50ml Self Stand (Bulk)	29.2	117.3	500
<b>P10487</b>	50ml Self Stand	29.2	117.3	500
<b>P10488</b>	50ml Self Stand (Sterile)	29.2	117.3	500

#### **1.4 Material:**



**Shiv Ganga Industrial Estate,**  
Plot no.1 Lakeshwari, Bhagwanpur,  
Roorkee – 247667 Uttarakhand, India  
**An ISO 9001, 13485, 14001,18385 & CE Certified**  
A Division of Abdos

**Regd. Office:**  
13A, Government Place (East)  
Kolkata – 700069, India

**P** : +91 (0) 1332 235277  
**E** : labtech@abdosindia.com  
**W**: www.abdoslifesciences.com  
**CIN** : U85100WB2008PTC123903  
**GST** : 05AAGCA8605A1Z7

# Technical Data Sheet

1.4.1 Tube - Polypropylene (PP) & Cap- High Density Polyethylene (HDPE) confirm US FDA 21CFR

1.4.2 Complies with the requirement of European Pharmacopoeia, Plastic Containers and closures for Pharmaceutical use.

1.4.3 The product has passed the USP testing including class VI tests. Materials has successful passed the biological tests according to ISO 10993 external communication devices for indirect blood contact for Prolonged period.

1.4.4 Free from Heavy metals, RoHS Free according to meets relevant requirements of EU regulations 1935/2004/CE ,2002/95/EC as amended, UE 10/2011 as amended. EU CONEG & it's following amendments. It is safe use in contact with food stuff, Pharmaceuticals & drinking water

1.4.5 Product is manufactured without the use of Raw Materials of animal origin. Therefore BSE/TSE issue does not concern this product to the best of our Knowledge. No materials containing latex or natural rubber are used in the Manufacturing, handling & packaging process for this product.

## 1.5 Color:

**Tube color-** Translucent (Natural)

**Cap color-** Natural

## 1.6 Graduation and writing surface:

1.6.1 Large Printed Graduation

1.6.2 White Patch for easy writing on the tube

## 1.7 Sterilization:

**Sterile catalog number** from this lot have been pre-sterilizes and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products Requirements for validation control- Radiation sterilization*. The product has been sterilized with Co-60 Gamma Radiation. SAL  $10^{-6}$

**1.8 Quality control:** Product are manufactured, inspected and found to be in compliance with product & quality specification requirements as documented in ISO 9001:2015 & 13485:2016 Quality Management System. Testing of various characteristics in accordance with the valid specification, inspection reviewed & signed for product release.

## 1.9 Other information

1.9.1 Recommended for single use only

1.9.2 Pack Size: 500pieces/case

1.9.3 Manufactured in Clean room ISO Class 8



**Shiv Ganga Industrial Estate,**  
Plot no.1 Lakeshwari, Bhagwanpur,  
Roorkee – 247667 Uttarakhand, India  
**An ISO 9001, 13485, 14001,18385 & CE Certified**  
A Division of Abdos

**Regd. Office:**  
13A, Government Place (East)  
Kolkata – 700069, India

**P** : +91 (0) 1332 235277  
**E** : labtech@abdosindia.com  
**W**: www.abdoslifesciences.com  
**CIN** : U85100WB2008PTC123903  
**GST** : 05AAGCA8605A1Z7

# Technical Data Sheet

## 2. Basic Features

**2.1 Temperature range:** Tube: -80°C to +121°C

**2.2 Autoclavability:** Yes, Autoclavable at 15 psi and 121°C for 20 minutes

**2.3 Centrifugation:** 18,000 RCF (15ml), 20,000 RCF (50ml conical) & 7,000 RCF (50ml self stand)

**2.4 Leak Test at 635 mm Hg:** No leakage


**2.5 SAL** 10<sup>-6</sup>

**2.6 Chemical resistance:** See website, [www.abdoslifesciences.com](http://www.abdoslifesciences.com)

**2.7 Trace Metals included in testing are:** Zinc, Iron, Boron, Copper, Manganese, Aluminum, Silicon, Nickel, Vanadium, Phosphorus, Cobalt, Chromium, Lithium, Selenium, Cadmium, Mercury, Arsenic & Lead ( 18)

## Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121°C for 20 minutes, 15 psi/1atm. Do not use cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50°C before removing product.

  
Quality Assurance Incharge  
Abdos Labtech Pvt. Ltd.



**ABDOS**

**Shiv Ganga Industrial Estate,**  
Plot no.1 Lakeshwari, Bhagwanpur,  
Roorkee – 247667 Uttarakhand, India

**An ISO 9001, 13485, 14001, 18385 & CE Certified**

A Division of Abdos

**Regd. Office:**

13A, Government Place (East)  
Kolkata – 700069, India

**P** : +91 (0) 1332 235277

**E** : [labtech@abdosindia.com](mailto:labtech@abdosindia.com)

**W:** [www.abdoslifesciences.com](http://www.abdoslifesciences.com)

**CIN** : U85100WB2008PTC123903

**GST** : 05AAGCA8605A1Z7