



Product Service

Attestation of Compliance

No. E8N 14 08 60707 058

Holder of Certificate: Eutech Instruments Pte Ltd
(Part of Thermo Fisher Scientific, Inc.)

Ayer Rajah Crescent
Blk 55 #04-16/24
Singapore 139949
SINGAPORE

Name of Object: Laboratory Equipment
(Water Analysis Instrument)

This Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

Test report no.: 7191091325-EEC14/01

Date, 2014-08-18

(Shuchen Li)



CE After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 2



Product Service

Attestation of Compliance
No. E8N 14 08 60707 058**Model(s):** pH 150, CON 150, pH 450, CON 450,
DO 450, PC 450, PD 450**Trade name:** EUTECH INSTRUMENTS**Description of Object:**
Input Rating : DC 9V, 1 A
or
Input Rating : 2 X AA (LR6)**Tested according to:** EN 61326-1:2013



Product Service

Attestation of Compliance

No. N5 14 08 60707 057

**Holder of Certificate: Eutech Instruments Pte Ltd
(Part of Thermo Fisher Scientific, Inc.)**

Ayer Rajah Crescent
Blk 55 #04-16/24
Singapore 139949
SINGAPORE

**Product: Laboratory Equipment
(Water Analysis Instrument)**

This Attestation of Compliance is issued on a voluntary basis for electrical equipment below the voltage limits of Low Voltage Directive 2006/95/EC (previous 73/23/EEC). The essential requirements are fulfilled accordingly. It refers to the particular sample submitted for testing and certification only. See also notes overleaf.

Test report no.: 7191091619-EEC14/NCH



Date, 2014-08-18

(Paul Chong Hing Tan)

CE Other relevant European directives have to be observed. If they require CE marking it may be affixed on the product after preparation of the necessary technical documentation as well as the declaration of conformity.

Page 1 of 2



Product Service

Attestation of Compliance
No. N5 14 08 60707 057

Model(s): pH 150, CON 150, pH 450, CON 450,
DO 450, PC 450, PD 450

Trade name: EUTECH INSTRUMENTS

Parameters: Input Rating : DC 9V, 1 A
or
Input Rating : 2 X AA (LR6)

Tested according to: EN 61010-1:2010
IEC 61010-1:2010