

multiSUB[®] Electrophoresis Systems MSMINI MSMIDI MSCHOICE MSCHOICEST MSMAXI

Instruction Manual

Catalogue No: MSMINI7, MSMINI10, MSMINIDUO
MSMIDI7, MSMIDI10, MSMIDIDUO
MSCHOICE7, MSCHOICE10, MSCHOICE15, MSCHOICETRIO
MSCHOICEST20, MSCHOICEST25
MSMAXI10, MSMAXI15, MSMAXI20, MSMAXIDUO, MSMAXI25




Table of Contents

	<u>Page</u>	
Section 1	Safety Information	3
1.1	Safety Precaution	3
Section 2	General Information	4
2.1	Packing Lists	4
2.2	Care and Maintenance	7
Section 3	Operating Instructions	8
3.1	Usage Guidance and restrictions	8
3.2	Setting up the Horizontal Gel Tanks	9
3.3	Gel Preparation	10
3.4	Gel Pouring	11
3.5	Running the Gel	13
3.6	Gel Staining and Viewing	14
3.7	Solutions	14
Section 4	References	15
Section 5	Trouble Shooting Guide	16
Section 6	Product Information	17
6.1	Catalogue numbers and product descriptions for multiSUB® units & accessories	17
6.2	Combs	19
6.3	Catalogue numbers and product descriptions for CSL Related Products	21
Section 7	Warranty	22

Section 1 Safety Information

1.1 Safety Precaution



When used correctly, these units pose no health risk. However, these units can deliver dangerous levels of electricity and are to be operated only by qualified personnel following the guidelines laid out in this instruction manual. Anyone intending to use this equipment should read the complete manual thoroughly. The unit must never be used without the safety lid correctly in position. The unit should not be used if there is any sign of damage to the external tank or lid.

These units comply with the following European directives:

2006/95/CE Low Voltage Directive and 2014/30/UE (official Title 2004/108/EC) EMC Electromagnetic Compatibility

By virtue of the following harmonised standards:

BS EN IEC 61010-1: 2010 Safety Testing of Lab Equipment

BS EN IEC 61326-1:2013 EMC Electro Magnetic Compatibility

Section 2 General Information

2.1 Packing Lists

MSMINI7, MSMINI10, MSMINIDUO

Units include tank, lid and electrodes and include the following accessories:

	Tray	Tray Dams	Combs	Loading Guides	Cables
MSMINI7	MS7-UV7 7 x 7cm (W x L)	MS7-UVDAM Pack of 2	2 x MS7-8-1 1mm, 8 sample	MS7-LG – Strips MS7-WP – Platform	CSL-CAB
MSMINI10	MS7-UV10 7 x 10cm (W x L)	MS7-UVDAM Pack of 2	2 x MS7-8-1 1mm, 8 sample	MS7-LG – Strips MS7-WP – Platform	CSL-CAB
MSMINIDUO	MS7-UV7 MS7-UV10	MS7-UVDAM Pack of 2	2 x MS7-8-1 1mm, 8 sample	MS7-LG – Strips MS7-WP – Platform	CSL-CAB

MSMIDI7, MSMIDI10, MSMIDIDUO

Units include tank, lid and electrodes and include the following accessories:

	Tray	Tray Dams	Combs	Loading Guides	Cables
MSMIDI7	MS10-UV7 10 x 7cm (W x L)	MS10- UVDAM Pack of 2	2 x MS10-16-1 1mm, 16 sample	MS10-LG – Strips MS10-WP – Platform	CSL-CAB
MSMIDI10	MS10-UV10 10 x 10cm (W x L)	MS10- UVDAM Pack of 2	2 x MS10-16-1 1mm, 16 sample	MS10-LG – Strips MS10-WP – Platform	CSL-CAB
MSMIDIDUO	MS10-UV7 MS10-UV10	MS10- UVDAM Pack of 2	2 x MS10-16-1 1mm, 16 sample	MS10-LG – Strips MS10-WP – Platform	CSL-CAB

MSCHOICE7, MSCHOICE10, MSCHOICE15, MSCHOICETRIO

Units include tank, lid and electrodes and include the following accessories:

	Tray	Tray Dams	Combs	Loading Guides	Cables
MSCHOICE7	MS15-UV7 15 x 7cm (W x L)	MS15- UVDAM Pack of 2	2 x MS15-20- 1 1mm, 20 sample	MS15-LG – Strips MS15-WP – Platform	CSL-CAB
MSCHOICE10	MS15-UV10 15 x 10cm (W x L)	MS15- UVDAM Pack of 2	2 x MS15-20- 1 1mm, 20 sample	MS15-LG – Strips MS15-WP – Platform	CSL-CAB
MSCHOICE15	MS15-UV15 15 x 15cm (W x L)	MS15- UVDAM Pack of 2	2 x MS15-20- 1 1mm, 20 sample	MS15-LG – Strips MS15-WP – Platform	CSL-CAB
MSCHOICETRIO	MS15-UV7 MS15-UV10 MS15-UV15	MS15- UVDAM Pack of 2	2 x MS15-20- 1 1mm, 20 sample	MS15-LG – Strips MS15-WP – Platform	CSL-CAB

MSCHOICEST20, MSCHOICEST25

Units include tank, lid and electrodes and include the following accessories:

	Tray	Tray Dams	Combs	Loading Guides	Cables
MSCHOICEST20	MS15- UVST20 15 x 20 cm (W x L)	MS15- UVDAM Pack of 2	4 x MS15- 28MC-1 1mm thick, 28 sample MC	MS15-LG – Strips MS15-WP – Platform	CSL-CAB
MSCHOICEST25	MS15- UVST25 15 x 25 cm (W x L)	MS15- UVDAM Pack of 2	4 x MS15- 28MC-1 1mm thick, 28 sample MC	MS15-LG – Strips MS15-WP – Platform	CSL-CAB

MSMAXI10, MSMAXI15, MSMAXI20, MSMAXIDUO, MSMAXI25

Units include tank, lid and electrodes and include the following accessories:

	Tray	Tray Dams	Combs	Loading Guides	Cables
MSMAXI10	MS20-UV10 20 x 10cm (W x L)	MS20-UVDAM Pack of 2	2 x MS20-20-1 1mm, 20 sample	MS20-LG – Strips MS20-WP – Platform	CSL-CAB
MSMAXI15	MS20-UV15 20 x 15cm (W x L)	MS20-UVDAM Pack of 2	2 x MS20-20-1 1mm, 20 sample	MS20-LG – Strips MS20-WP – Platform	CSL-CAB
MSMAXI20	MS20-UV20 20 x 20cm (W x L)	MS20-UVDAM Pack of 2	2 x MS20-20-1 1mm, 20 sample	MS20-LG – Strips MS20-WP – Platform	CSL-CAB
MSMAXIDUO	MS20-UV10 MS20-UV20	MS20-UVDAM Pack of 2	2 x MS20-20-1 1mm, 20 sample	MS20-LG – Strips MS20-WP – Platform	CSL-CAB
MSMAXI25	MS20-UV25 20 x 25cm (W x L)	MS20-UVDAM Pack of 2	2 x MS20-20-1 1mm, 20 sample	MS20-LG – Strips MS20-WP – Platform	CSL-CAB

Packing list checked by _____

Date _____

The packing lists should be referred to as soon as the units are received to ensure that all components have been included. The unit should be checked for damage when received.

Please contact your supplier if there are any problems or missing items.

2.2 Care and Maintenance

Cleaning Horizontal Units

Units are best cleaned using warm water and a mild detergent. **Water at temperatures above 60°C can cause damage to the unit and components.**

The tank should be thoroughly rinsed with warm water or distilled water to prevent buildup of salts but care should be taken not to damage the enclosed electrode and vigorous cleaning is not necessary or advised.

Air drying is preferably before use.

The units should only be cleaned with the following:

Warm water with a mild concentration of soap or other mild detergent.

Compatible detergents include dishwashing liquid, Hexane and Aliphatic hydrocarbons

The units should not be left to in detergents for more than 30 minutes.

The units should never come into contact with the following cleaning agents, these will cause irreversible and accumulative damage:

Acetone, Phenol, Chloroform, Carbon tetrachloride, Methanol, Ethanol, Isopropyl alcohol

Alkalis.

Rnase Decontamination

This can be performed using the following protocol:

Clean the units with a mild detergent as described above.

Wash with 3% hydrogen peroxide (H₂O₂) for 10 minutes.

Rinsed with 0.1% DEPC- (diethyl pyro carbonate) treated distilled water,

Caution: DEPC is a suspected carcinogen. Always take the necessary precautions when using.

RNaseZAP™ (Ambion) can also be used. Please consult the instructions for use with acrylic gel tanks.

Section 3 Operating Instructions

Further information (including videos) regarding setting up and running the multiSUB® units can be found at www.cleaverscientific.com

3.1 Usage Guidance and restrictions

- Maximum altitude 2,000m.
- Temperature range between 4°C and 65°C.
- Maximum relative humidity 80% for temperatures up to 31°C decreasing linearly to 50% relative humidity at 40°C.
- Not for outdoor Use.

This apparatus is rated POLLUTION DEGREE 2 in accordance with IEC 664.

POLLUTION DEGREE 2, states that: "Normally only non-conductive pollution occurs.

Occasionally, however, a temporary conductivity caused by condensation must be expected".

3.2 Setting up the Horizontal Gel Tanks

Instructions for fitting Electrode Cables

1. Note the position of the lid on the unit. This shows the correct polarity and the correct orientation of the cables, black is negative and red positive.
2. Remove the lid from the unit. Note if the lid is not removed, fitting the cables may result in un-tightening of the gold plug and damage to the electrode.
3. Screw the cables into the tapped holes as fully as possible so that there is no gap between the lid and the leading edge of the cable fitting.
4. Refit the lid.

Instructions for fitting Loading Guides

These can be fitted to enhance visibility of the wells if desired. They can be fitted to the white vinyl platform sheet or to the unit itself.

1. Seat the tray in the unit and note the position of the comb grooves. The samples run black to red but the trays can be used frontward or backwards so ensure that the comb grooves closest to the black electrode are marked.
2. Remove the tray.
3. Peel the back off of the loading guide and carefully apply the loading guide directly to the gel platform.

The unit is now ready to be used.

3.3 Gel Preparation

1. Table 1 below shows the volume of agarose solution required to make the desired agarose gel for each unit tray size. For a standard 0.7% agarose gel, add 0.7 grams of agarose to 100 ml of 1x TAE or TBE solution. The same 1 x solution should be used in the tank buffer solution.

Table 1

multiSUB® Mini	Tray	7 x 7 cm	7 x 10 cm	
	Gel volume for a 5mm thick gel	25 mL	35 mL	
multiSUB® Midi	Tray	10 x 7 cm	10 x 10 cm	
	Gel volume for a 5mm thick gel	35 mL	50 mL	
multiSUB® Choice	Tray	15 x 7 cm	15 x 10 cm	15 x 15 cm
	Gel volume for a 5mm thick gel	52.5 mL	75 mL	112.5 mL
multiSUB® Choice STRETCH	Tray	15 x 20 cm	15 x 25 cm	
	Gel volume for a 5mm thick gel	150 mL	187.5 mL	
multiSUB® Maxi	Tray	20 x 10 cm	20 x 15 cm	20 x 20 cm
	Gel volume for a 5mm thick gel	100 mL	150 mL	200 mL

2. Add the agarose powder to a conical flask.
3. Add the appropriate amount of 1x TAE or TBE solution from the table above. To prevent evaporation during the dissolving steps below, the conical flask should be covered with parafilm.
4. Dissolve the agarose powder by heating the agarose either on a magnetic hot plate with stirring bar or in a microwave oven. If using the microwave method, the microwave should be set at around a 400 watt or medium setting and the flask swirled every minute. The solution should be heated until all crystals are dissolved. This is best viewed against a light background. Crystals appear as translucent crystals. These will interfere with sample migration if not completely dissolved.

The gel must be cooled to between 50°C and 60°C degrees before pouring.

3.4 Gel Pouring

The CSL Multi sub range of units allows three different methods of gel casting:

Casting Dams

Flexicaster

Traditional Tape

Using trays with Casting Dams

1. To fit the casting dams, place one casting dam on the bench with the groove facing upwards (see figure 1). Push the edge of the tray down firmly into the groove (see figure 2). Repeat this for the other side (figure 3). The dams should be fitted so that there is no gap between the sides of the tray and the groove in the dams. This will ensure that there is no possibility of gel leakage.
2. Place the comb(s) in the grooves. Each tray has more than one comb groove so that multiple combs can be used. Using multiple combs increases sample number available per gel but decreases run length and care must be taken to ensure that samples from the first wells do not migrate into the lanes of the second comb wells.
3. Pour in the agarose carefully so as not to generate bubbles. Any bubbles that do occur can be smoothed to the edge of the gel and dispersed using a pipette tip.
4. Allow the agarose to set, ensuring that the gel remains undisturbed.
5. Carefully remove the gel casting gates and comb and transfer the gel including tray to the main tank.

Figures 1 – 3 Assembling the casting dam to the tray



Figure 1



Figure 2



Figure 3

Using The Flexicaster

See figures 4 and 5 for Flexicaster MS7-FC and MS20-FC

1. Level the Flexicaster base by adjusting the feet so that the bubble is exactly central.
2. Insert the desired length tray into the Flexicaster such that one end of the tray is pushed up and seals against the silicone mat of the permanent end of the Flexicaster.
3. Position the movable end of the Flexicaster so that the silicone mat is pushed against the other end of the tray.
4. Turn the cam so that the silicone mat tightly seals against the side of the tray. Pour in the agarose carefully so as not to generate bubbles. Any bubbles that do occur can be smoothed to the edge of the gel and dispersed using a pipette tip.
5. Allow the agarose to set, ensuring that the gel remains undisturbed.
6. Carefully remove the gel casting gates and comb and transfer the gel including tray to the main tank.



Figure 4 MS7-FC



Figure 5 MS20-FC

Using Traditional tape method

1. Autoclave or plastic backed general tape should be used. A length 5cm longer than the width of each end of the tray should be cut. One length should be placed over one end of the tray and stuck 1cm in from the tray edge. This should then be folded and the edges sealed securely. Repeat for the other end and place onto a level surface for gel pouring.
2. Place the comb(s) in the grooves. Each tray has more than one comb groove so that multiple combs can be used. Using multiple combs increases sample number available per gel but decreases run length and care

must be taken to ensure that samples from the first wells do not migrate into the lanes of the second comb wells.

3. Pour in the agarose carefully so as not to generate bubbles. Any bubbles that do occur can be smoothed to the edge of the gel and dispersed using a pipette tip.
4. Allow the agarose to set, ensuring that the gel remains undisturbed.
5. Carefully remove the gel casting gates and comb and transfer the gel including tray to the main tank.

3.5 Running the Gel

1. Mix the sample to be loaded with sample buffer – see solutions for common sample buffers. Usually 3ul of sample buffer is adequate but less may be used with sample volumes of less than 10ul.
2. Fill the unit with buffer until the gel is just flooded with buffer. This will give the fastest resolution times. For enhanced quality of resolution of sample, fill the unit to 5mm above the gel.
3. Load the samples into the wells using pipettes. Multi-channel pipettes can be used for loading samples with MC compatible combs, see listing in accessories for identification of these.
4. Carefully place the lid on the tank and connect to a power supply.
5. Typically gels are run at between 90 and 150 volts. However, maximum voltages are indicated on the serial badge of each unit. It should be noted that higher voltages generally give faster but poorer quality sample resolution.

3.6 Gel Staining and Viewing

The Multi Sub trays allow staining to be performed without removing the gel from the tray if this is preferred.

1. Transfer the gel to a vessel containing the appropriate volume of 0.5 µg/ml ethidium bromide stain for 15–30 minutes, see solutions for stock stain concentration and adjust to the volume used accordingly. The entire gel should be covered.

NOTE: Ethidium bromide is a suspected carcinogen and the necessary safety precautions should be undertaken.

2. De-stain the gel for 10–30 minutes in distilled water again ensuring the gel is completely immersed.
3. Rinse the gel twice for a couple of seconds with distilled water.
4. Transfer the gel to a UV Transilluminator.
5. The samples will often appear as brighter, clearer bands when photographed or viewed using a gel documentation system. However if the gel bands are too faint then the staining procedure should be adjusted so that there is less de-staining. If there is too much background then the staining procedure should be adjusted so that there is more de-staining.

3.7 Solutions

1x TAE 40 mM tris (pH 7.6), 20 mM acetic acid, 1 mM EDTA.

50x (1L) dissolve in 750 ml distilled water:

242 g tris base (FW = 121)

57.1 ml glacial acetic acid

100 ml 0.5 M EDTA (pH 8.0).

Fill to 1 litre with distilled water.

1x TBE 89 mM tris (pH 7.6), 89 mM boric acid, 2 mM EDTA

10x (1L) dissolve in 750 ml distilled water:

108 g tris base (FW = 121)

55 g boric acid (FW = 61.8)

40 ml 0.5 M EDTA (pH 8.0)

Fill to 1 litre with distilled water.

Sample Loading Dye

10x sample buffer stock consists of 50% glycerol, 0.25% bromophenol blue, and

0.25% xylene cyanole FF in 1x TAE buffer. Only 1–10 ml of the 10x loading dye should be prepared.

Ethidium Bromide Solution

Add 10 mg of Ethidium Bromide to 1 ml distilled water.

Section 4 References

1. Sambrook, Fritsch, and Maniatis, **Molecular Cloning A Laboratory Manual**, Second Edition, Cold Spring Harbor Laboratory Press, 1989.
2. **Current Protocols in Molecular Biology**, Greene Publishing Associates and Wiley-Interscience, 1989.

Section 5 Troubleshooting Guide

Problem	Cause	Solution
Bands sharp but not enough bands seen	Gel agarose percentage too high Incomplete digestion	Decrease agarose percentage. Review enzyme activity, digest further.
Band smearing and streaking	Agarose has improper endosmosis Salt concentration in sample too high Excessive power and heating Sample spilled out of well Incomplete digestion, nuclease contamination, bad enzyme Sample wells cast through the gel. Sample leaks along bottom of running surface	Consult Cleaver Scientific about agarose. Reduce salt concentration to $\leq 0.1M$. Reduce voltage. See electrophoresis instructions. Apply sample carefully. Increase gel thickness for large sample volumes. Adjust comb height. Heat sample. Review enzyme activity. Digest sample further. Comb should be placed to 1 to 2 mm above the base of the running surface.
Curved line or distortion of bands	Bubbles in sample wells	Remove bubbles prior to electrophoresis.
Curved bands, smiles	Sample overload	Reduce load.
Differential relative mobilities	Sample spilled out of wells Unit not leveled	Samples should have proper density. Apply carefully. Level unit. Use a steady work bench.
Gels crack	Too high voltage gradient, especially with low melting temperature agarose or low gel strength gels	Reduce voltage. Run gel at lower temperature.
High MW bands sharp; low MW bands smeared	Gel agarose percentage too low	Increase agarose percentage. Switch to polyacrylamide.
Ragged bands	Sample density incorrect Sample well deformed Excessive power or heating	See sample application instructions. Carefully remove comb, especially from soft gels. Make sure gel has solidified. Cooling soft gels aids in comb removal. Reduce voltage. See electrophoresis instructions.
Slanted lanes (bands)	Gel not fully solidified Comb warped or at an angle	Gel to solidify for at least 30-45min. Check alignment of comb.

Section 6 Product Information

6.1 Catalogue numbers and product descriptions for multiSUB® units & accessories

Catalogue No.	Product description
MSMINI7	multiSUB® Mini, 7 x 7cm UV Tray, 2 x 8 sample combs, loading guides and dams
MSMINI10	multiSUB® Mini, 7 x 10cm UV Tray, 2 x 8 sample combs, loading guides and dams
MSMINIDUO	multiSUB® Mini, 7 x 7cm & 7 x 10cm UV Tray, 2 x 8 sample combs, loading guides and dams
MS7-UV7	7 x 7cm UV Tray
MS7-UV10	7 x 10cm UV Tray
MS7-PE	Positive Electrode
MS7-NE	Negative Electrode
MS7-UVDAM	Casting Dams
MS7/10-FC	multiSUB® Mini/Midi Flexi caster
MS7-LG	Adhesive Loading Guides
MS7-WP	Viewing Platform
MSMINICP	Cool-pack and Platform
MSMINIBSB	Buffer Saver Blocks, pk/2 saves 100ml of buffer
MS7-UVS	7cm UV Gel Scoop
CSL-CAB	Electrophoresis cable (Black & Red)
MSMIDI7	multiSUB® Midi, 10 x 7cm UV Tray, 2 x 16 sample combs, loading guides and dams
MSMIDI10	multiSUB® Midi, 10 x 10cm UV Tray, 2 x 16 sample combs, loading guides and dams
MSMIDIDUO	multiSUB® Midi, 10 x 7cm & 10 x 10cm UV Tray, 2 x 16 sample combs, loading guides and dams
MS10-UV7	10 x 7cm UV Tray
MS10-UV10	10 x 10cm UV Tray
MS10-PE	Positive Electrode
MS10-NE	Negative Electrode
MS10-UVDAM	Casting Dams
MS7/10-FC	multiSUB® Mini/Midi Flexi caster
MS10-LG	Adhesive Loading Guides
MS10-WP	Viewing Platform
MSMIDICP	Cool-pack and Platform
MSMIDIBSB	Buffer Saver Blocks, pk/2 saves 100ml of buffer
MS10-UVS	10cm UV Gel Scoop
CSL-CAB	Electrophoresis cable (Black & Red)

MSCHOICE7	multiSUB®™ Choice 15 x 7cm UV Tray, 2 x 20 sample combs, loading guides and dams
MSCHOICE10	multiSUB®™ Choice 15 x 10cm UV Tray, 2 x 20 sample combs, loading guides and dams
MSCHOICE15	multiSUB®™ Choice 15 x 15cm UV Tray, 2 x 20 sample combs, loading guides and dams
MSCHOICETRIO	multiSUB®™ Choice Trio 15 x 7,10 & 15cm UV Tray, 2 x 20 sample combs, loading guides and dams
MSCHOICETRIO15	multiSUB®™ Choice Trio 15 3 x 15cm UV tray, 2x 20 sample combs, loading guides and dams
MSCHOICEST20	multiSUB®™ Choice Stretch 20 15 x 20cm UV tray, 4 x 28 sample combs, loading guides and dams
MSCHOICEST25	multiSUB®™ Choice Stretch 25 15 x 25cm UV tray, 4 x 28 sample combs, loading guides and dams
MS15-UV7	15 x 7cm UV Tray
MS15-UV10	15 x 10cm UV Tray
MS15-UV15	15 x 15cm UV Tray
MS15-UVST20	15 x 20cm UV Tray
MS15-UVS	15cm UV Gel Scoop
MS15-UVDAM	Casting Dams
MS15-PE	Positive Electrode
MS15-NE	Negative Electrode
MSMS15-LG	Adhesive Loading Guides
MS15-WP	Viewing Platform
MSCHOICECP	Cool-Pack and Platform
MS15/20-FC	multiSUB® Choice/ Maxi Flexi caster
MSMAXI10	20 x 10cm UV Tray, 2 x 20 sample, 1mm thick combs, casting dams, loading guides
MSMAXI15	20 x 15cm UV Tray, 2 x 20 sample, 1mm thick combs, casting dams, loading guides
MSMAXI20	20 x 20cm UV Tray, 2 x 20 sample, 1mm thick combs, casting dams, loading guides
MSMAXIDUO	20 x 10 & 20 x 20cm UV Tray, 2 x 20 sample, 1mm thick combs, casting dams, loading guides
MSMAXI25	20 x 25cm UV Tray, 2 x 20 sample, 1mm thick combs, casting dams, loading guides
MS20-UV10	20 x 10cm UV Tray
MS20-UV15	20 x 15cm UV Tray
MS20-UV20	20 x 20cm UV Tray
MS20-UV25	20 x 25cm UV Tray
MS20-UVDAM	Casting Dams
MS15/20-FC	multiSUB Choice/Maxi Flexi caster
MS20-LG	Adhesive Loading Guides
MS20-WP	Viewing Platform
MS20-PE	Positive Electrode
MS20-NE	Negative Electrode
MS20-UVS	20cm UV Gel Scoop
MSMAXICP	Cool-Pack and Platform
MSMAXIBSB	Buffer Saver Blocks pk/2 saves 450ml of buffer

6.2 Combs – MC Denotes Multi Channel Pipette compatible

multiSUB® Mini 0.75mm		multiSUB® Midi 0.75mm		multiSUB® Choice 0.75mm		Multi Sub Maxi 0.75mm	
No. of Samples	Part Number	No. of Samples	Part Number	No. of Samples	Part Number	No. of Samples	Part Number
1	MS7-1-0.75	1	MS10-1-0.75	1	MS15-1-0.75	1	MS20-1-0.75
2	MS7-2-0.75	2	MS10-2-0.75	2	MS15-2-0.75	2	MS20-2-0.75
4	MS7-4-0.75	4	MS10-4-0.75	4	MS15-4-0.75	4	MS20-4-0.75
8	MS7-8MC-0.75 MS7-8-0.75	8	MS10-8-0.75	10	MS15-10-0.75 MS15-10MC-0.75	10	MS20-10-0.75
10	MS7-10-0.75	10	MS10MC-10-0.75	12	MS15-12-0.75	16	MS20-16-0.75
12	MS7-12-0.75	12	MS10-12-0.75	14	MS15-14MC-0.75	20	MS20-20MC-0.75
16	MS7-16-0.75	16	MS10-16-0.75	16	MS15-16-0.75	25	MS20-25-0.75
		20	MS10-20MC-0.75	18	MS15-18MC-0.75	30	MS20-30-0.75
				20	MS15-20-0.75	36	MS20-36-0.75
				28	MS15-28MC-0.75	40	MS20-40MC-0.75
				30	MS15-30-0.75		

multiSUB® Mini 1mm		multiSUB® Midi 1mm		multiSUB® Choice 1mm		Multi Sub Maxi 1mm	
No. of Samples	Part Number	No. of Samples	Part Number	No. of Samples	Part Number	No. of Samples	Part Number
1	MS7-1-1	1	MS10-1-1	1	MS15-1-1	1	MS20-1-1
2	MS7-2-1	2	MS10-2-1	2	MS15-2-1	2	MS20-2-1
4	MS7-4-1	4	MS10-4-1	4	MS15-4-1	4	MS20-4-1
8	MS7-8MC-1 MS7-8-1	8	MS10-8-1	10	MS15-10-1 MS15-10MC-1	10	MS20-10-1
10	MS7-10-1	10	MS10MC-10-1	12	MS15-12-1	16	MS20-16-1
12	MS7-12-1	12	MS10-12-1	14	MS15-14MC-1	20	MS20-20MC-1
16	MS7-16-1	16	MS10-16-1	16	MS15-16-1	25	MS20-25-1
		20	MS10-20MC-1	18	MS15-18MC-1	30	MS20-30-1
				20	MS15-20-1	36	MS20-36-1
				28	MS15-28MC-1	40	MS20-40MC-1
				30	MS15-30-1		

multiSUB® Mini 1.5mm		multiSUB® Midi 1.5mm		multiSUB® Choice 1.5mm		Multi Sub Maxi 1.5mm	
No. of Samples	Part Number	No. of Samples	Part Number	No. of Samples	Part Number	No. of Samples	Part Number
1	MS7-1-1.5	1	MS10-1-1.5	1	MS15-1-1.5	1	MS20-1-1.5
2	MS7-2-1.5	2	MS10-2-1.5	2	MS15-2-1.5	2	MS20-2-1.5
4	MS7-4-1.5	4	MS10-4-1.5	4	MS15-4-1.5	4	MS20-4-1.5
8	MS7-8MC-1.5 MS7-8-1.5	8	MS10-8-1.5	10	MS15-10-1.5 MS15-10MC-1.5	10	MS20-10-1.5
10	MS7-10-1.5	10	MS10MC-10-1.5	12	MS15-12-1.5	16	MS20-16-1.5
12	MS7-12-1.5	12	MS10-12-1.5	14	MS15-14MC-1.5	20	MS20-20MC-1.5
16	MS7-16-1.5	16	MS10-16-1.5	16	MS15-16-1.5	25	MS20-25-1.5
		20	MS10-20MC-1.5	18	MS15-18MC-1.5	30	MS20-30-1.5
				20	MS15-20-1.5	36	MS20-36-1.5
				28	MS15-28MC-1.5	40	MS20-40MC-1.5
				30	MS15-30-1.5		

multiSUB® Mini 2mm		multiSUB® Midi 2mm		multiSUB® Choice 2mm		Multi Sub Maxi 1.75mm	
No. of Samples	Part Number	No. of Samples	Part Number	No. of Samples	Part Number	No. of Samples	Part Number
1	MS7-1-2	1	MS10-1-2	1	MS15-1-2	1	MS20-1-2
2	MS7-2-2	2	MS10-2-2	2	MS15-2-2	2	MS20-2-2
4	MS7-4-2	4	MS10-4-2	4	MS15-4-2	4	MS20-4-2
8	MS7-8MC-2 MS7-8-2	8	MS10-8-2	10	MS15-10-2 MS15-10MC-2	10	MS20-10-2
10	MS7-10-2	10	MS10MC-10-2	12	MS15-12-2	16	MS20-16-2
12	MS7-12-2	12	MS10-12-2	14	MS15-14MC-2	20	MS20-20MC-2
16	MS7-16-2	16	MS10-16-2	16	MS15-16-2	25	MS20-25-2
		20	MS10-20MC-2	18	MS15-18MC-2	30	MS20-30-2
				20	MS15-20-2	36	MS20-36-2
				28	MS15-28MC-2		
				30	MS15-30-2		

6.3 Catalogue numbers and product descriptions for CSL Related Products

Catalogue No.	Product description
nanoPAC-300	Mini Power supply, 300V, 400mA, 60W -100 -240VAC
nanoPAC-500	Mini Power supply, 500V, 400mA, 120W -100 -240VAC
CS300V	omniPAC MIDI Power Supply, 300V, 700mA, 150 – 100 -240VAC
OmniDOCIPROSAFE	OMNIDOC-i plus Blue LED Epi-illumination Module (OMNIDOC-BL), and 520, 560 & 580nm filters (OMNIDOC-SYBR, -AF560 & -AF580); and White Light Table (OMNIDOC-WLT). Requires a PC, laptop or tablet with Wi-Fi connection
CSL-AG5	Agarose Powder 5g, Low EEO
CSL-AG100	Agarose Powder 100g, Low EEO
CSL-AG500	Agarose Powder 500g, Low EEO
CSL-AG1000	Agarose Powder 1000g, Low EEO (2x500g bottles)
CSL-AG2000	Agarose Powder 2000g, Low EEO (4x500g bottles)
CSL-AG5000	Agarose Powder 5000g, Low EEO (10x500g bottles)
CSL-AG10KG	Agarose Powder 10Kg, Low EEO (20x500g bottles)
CSL-LMA5	Agarose Powder 5g, Low Melting Point
CSL-LMA50	Agarose Powder 50g, Low Melting Point
CSL-LMA100	Agarose Powder 100g, Low Melting Point
CSL-HRA100	Agarose Powder 100g, High Resolution
CSL-HRA500	Agarose Powder 500g, High Resolution
CSL-AGTAB	Agarose Tablet 100g, Low EEO (200 x 0.5g tablets, supplies as 20 blister packs of 10 x 0.5g tablets)
CSL-RUNSAFE	CSL-RUNSAFE - Package: 1 ml/vial
CSL-TBEP	Powdered Tris-Borate-EDTA Running Buffer- 10 sachets (1litre/pk)
TBE10X1L	Cleaver Buffer Tris-Borate-EDTA Running Buffer- 10 x 1L
TBE10X5	Cleaver Buffer Tris-Borate-EDTA Running Buffer- 10 x 5L
TAE50X1L	Cleaver Buffer Tris-Borate-EDTA Running Buffer- 50 x 1L
TAE50X5L	Cleaver Buffer Tris-Borate-EDTA Running Buffer- 50 x 5L
CSL-MDNA-100BH	100bp DNA ladder, 100 – 300bp, 1 x 500µl vial
CSL-MDNA-100BP	100bp DNA ladder, 100 – 1500bp, 1 x 500µl vial
CSL-MDNA-1KB	1Kb DNA ladder, 250 – 10Kb, 1 x 500µl vial
CSL-MDNA-50BP	50bp DNA ladder, 50 – 1500bp, 1 x 500µl vial
CSL-MDNA-BR	Broad Range DNA ladder, 100bp – 10Kb, 1 x 500µl vial
CSL-MDNA-HR	High Range DNA ladder, 250bp – 25Kb, 1 x 500µl vial
CSL-LOADDYE	10x Bromophenol Blue Loading Dye, 1mL
CSL-LOADDYE10	10x Bromophenol Blue Loading Dye, 10mL
SAFEVIEW	BLUE Light Transilluminator 21 x 21cm
CSLUVTS312	UV Transilluminator, small 21 x 21 cm, 312nm
CSL-GELX4	4mm x 1mm, Gel Cutting Tips, 250/ bag
CSL-GELX4RACK	4mm x 1mm, Gel Cutting Tips, 5 racks of 48
CSL-GELX6.5	6.5mm x 1mm, Gel Cutting Tips, 250/ bag
CSL-GELX6.5RACK	6.5mm x 1mm, Gel Cutting Tips, 5 racks of 48
CSLQSPIN	Mini Centrifuge complete with 1.5/2.0 ml rotor, strip tube rotor, 0.5 and 0.4 ml adapters, 230V, Purple lid
CV20	Cleaver Pipette - Volume; 2 - 20µl



Section 7 Warranty

The Cleaver Scientific Ltd. (CSL) multiSUB® Horizontal Electrophoresis units have a warranty against manufacturing and material faults of twelve months from date of customer receipt.

If any defects occur during this warranty period, CSL will repair or replace the defective parts free of charge.

This warranty does not cover defects occurring by accident or misuse or defects caused by improper operation.

Units where repair or modification has been performed by anyone other than CSL or an appointed distributor or representative are no longer under warranty from the time the unit was modified.

Units which have accessories or repaired parts not supplied by CSL or its associated distributors have invalidated warranty.

CSL cannot repair or replace free of charge units where improper solutions or chemicals have been used. For a list of these please see the Care and Maintenance subsection.

If a problem does occur then please contact your supplier or CSL:

Cleaver Scientific Ltd.
Unit 41
Somers Road Industrial Estate
Rugby
Warwickshire
CV22 7DH
Tel: +44 (0)1788 565300
Fax: +44 (0)1788 552822
Email: info@cleaverscientific.com

Record the following for your records:

Model _____
Catalogue No. _____
Date of Delivery _____
Warranty Period _____
Serial No. _____
Invoice No. _____
Purchase Order No. _____