

Solstice

CERTIFICATE OF ANALYSIS

Honeywell Burdick & Jackson®

1953 South Harvey Street
Muskegon, MI 49442
Phone: (800) 368-0050
Fax: (231) 728-8226
lab.honeywell.com

Brand: Research Chemicals - Riedel-de Haën
Product: 34480-4X4L
Lot No.: EL059
Production Date: 10-Sep-2025
Best Before: 31-Aug-2027

Acetone, CHROMASOLV™, for pesticide residue analysis, >99.8%

Parameter	Specification		Result	Units
	Min.	Max.		
Assay by GC Analysis	99.8		99.96	%
Non-volatile Matter		0.0005	<0.0005	%
Water by Karl Fischer Titration		0.50	0.22	%
Suitability for Residue Analysis (ECD) as Heptachlor Epoxide		5	<5	ng/L

Honeywell

Solstice Advanced Materials

Quality Control Approval

