

**Honeywell**

## CERTIFICATE OF ANALYSIS

Honeywell Specialty Chemicals Seelze GmbH  
 Wunstorferstrasse 40  
 Seelze, 30926  
 Telefon: +49 5137 999-0  
[lab.honeywell.com](http://lab.honeywell.com)

Brand: Honeywell Riedel-de Haën  
 Product: 34860  
 Lot No: P058E  
 Production date: 27.Feb.2025  
 Expiry Date: 05.Aug.2029

Methanol CHROMASOLV, Reag. JP, for HPLC, >=99,9%

Parameter	Specification	Units	Result	Units
assay (GC)	min. 99.90	%	99.99	%
non-volatile matter	max. 0.0005	%	<0.0005	%
water (Karl Fischer)	max. 0.03	%	<0.001	%
absorbance at 400 nm	max. 0.01		<0.01	
absorbance 280 nm	max. 0.01		<0.01	
absorbance at 260 nm	max. 0.04		<0.04	
absorbance at 254 nm	max. 0.02		<0.02	
absorbance at 240 nm	max. 0.07		<0.07	
absorbance at 235 nm	max. 0.10		<0.10	
absorbance at 230 nm	max. 0.15		<0.15	
absorbance at 220 nm	max. 0.30		<0.30	
absorbance at 210 nm	max. 0.60		<0.60	
absorbance at 205 nm	max. 1.00		<1.00	
absorbance (JP)(Val.Ref)	complying		complying	
fluorescence (chinin) at 254 nm	max. 1	ppb	<1	ppb
fluorescence (chinin) at 365 nm	max. 1	ppb	<1	ppb
APHA	max. 10		<10	

filtered via 0,2µm pore size filter

QC Release Date: 14.Apr.2025

The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.

We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.

Particular properties of the products or the suitability for a particular area of application are not assured.

We guarantee a proper quality within our General Conditions of Sales.

Honeywell

Quality Management

This document was produced electronically and is valid without a signature