Biosigma S.p.A.
a Dominique Dutscher Company
Via Valletta, 6 | 30010 Cantarana di Cona (VE), Italy | Tel. ++39 0426 302224 (r.a.) |
Fax ++39 0426 302228 | SMS ++39 348 4077376 |
E-mail info@biosigmaeu.com | http://www.biosigma.com |

M521D

Ultimo aggiornamento
01/06/2023

SCHEDA TECNICA PRODOTTO / TECHNICAL DATA SHEET

BSP060
Provetta conica 10 ml
Conical test tube 10 ml
Conf. 2.000 pz/pcs.

BSP061
Provetta conica 10 ml - etichettata
Conical test tube 10 ml - with label
Conf. 2.000 pz/pcs



Altezza provetta	103 mm	***Tube height***
Diametro provetta	16 mm	***Tube diameter***
Volume	10 ml	***Volume***
Gradazione	SI / YES	***Scale area***
tappata	NO	***With cap***
Materiale	PP	***Raw material***
Sterile	NO	***Sterile***
Confezionamento	2000 pz/pcs	***Package***

Destinato esclusivamente ad uso professionale

- **Per usi generici di laboratorio**
- In caso di fuoriuscita del contenuto usare i guanti: pericolo di contaminazione
- Non avvicinare il dispositivo alla fiamma o a fonti di calore che lo potrebbero danneggiare
- Non utilizzare il prodotto scaduto
- Non variare la destinazione d'uso
- Prodotto non adatto ai bambini
- Conservare in luogo asciutto, temperatura min - 10 °C max + 50 °C
- Smaltimento: utilizzare gli appositi D.P.I. e smaltire secondo la normativa vigente

for professional use only

- ***For general labware use***
- *Use gloves: contamination risk if contents leaks*
- *Keep out of flame or heat sources which might damage the product*
- *Do not use after expiry date*
- *Do not vary the intended purpose of the product*
- *Keep out of reach of children*
- *Store in dry place, temperature range: min. - 10°C max 50°C*
- *Disposal: use appropriate personal protective equipment and act according to applicable regulations*

DEFINIZIONE DEL FABBRICANTE / *MANUFACTURER DEFINITION*

BIOSIGMA s.r.l.
Via Valletta, 6 – 30010 - Cantarana di Cona (VE) - ITALY
Tel. +39.0426.302224 - Fax +39.0426.302228
E-mail: info@biosigma.com - www. biosigma.it ; www. biosigma.com

DOMINIQUE DUTSCHER SAS