
Biosigma SpA
a Dominique Dutscher Company
Via Valletta, 6 | 30010 Cantarana di Cona (VE), Italy | Tel. ++39 0426 302224 (r.a.) |
Fax ++39 0426 302228 E-mail info@biosigmaeu.com | http://www.biosigma.com |

SCHEDA TECNICA PRODOTTO / TECHNICAL DATA SHEET

BSP072
Provetta cilindrica 10 ml
Cylindrical test tube 10 ml
Conf. 2.000 pz/pcs.

BSP073
Provetta cilindrica 10 ml - **etichettata**
*Cylindrical test tube 10 ml - **with label***
Conf. 2.000 pz/pcs





Altezza provetta	100 mm	***Tube height***
Diametro provetta	16 mm	***Tube diameter***
Volume	10 ml	***Volume***
Gradazione	2,5 – 5 – 10 ml	**Scale area**
Materiale	PS	**Raw material**
Sterile	NO	***Sterile***
Confezionamento	2000 pz/pcs	***Package***
velocità in centrifuga (max 30 minuti)	2.000 rpm	***centrifuge rotation speed (max 30 minutes)***

PER USO GENERICO DI LABORATORIO
Destinato esclusivamente ad uso professionale

- In caso di fuoriuscita del contenuto usare i guanti: pericolo di contaminazione
- Non avvicinare il dispositivo alla fiamma o a fonti di calore che lo potrebbero danneggiare
- Non utilizzare il prodotto scaduto
- Non variare la destinazione d'uso
- Prodotto non adatto ai bambini
- Conservare in luogo asciutto, temperatura min - 10 °C max + 50 °C
- Smaltimento: utilizzare gli appositi D.P.I. e smaltire secondo la normativa vigente

FOR GENERAL LABORATORY USE
for professional use only

- *Use gloves: contamination risk if contents leaks*
- *Keep out of flame or heat sources which might damage the product*
- *Do not use after expiry date*
- *Do not vary the intended purpose of the product*
- *Keep out of reach of children*
- *Store in dry place, temperature range: min. -10°C max 50°C*
- *Disposal: use appropriate personal protective equipment and act according to applicable regulations*

BIOSIGMA SpA Via Valletta, 6 – 30010 - Cantarana di Cona (VE) - ITALY
Tel. +39.0426.302224 - Fax +39.0426.302228
E-mail: info@biosigma.com - www. biosigma.it ; www. biosigma.com

DOMINIQUE DUTSCHER SAS