




# LEAK TEST

M611-LT

Prima emissione  
First emission 24/10/2013



Date 28/05/2019

REF CL5ARBIPSTS/BI

LOT 19P1050900

Progressive Number 1

Test Number 1

Pressure Fill Set 1000,0

Delta Pressure Set 0,2

Settling Time Set 8,0

Test Time Set 2,0

Test Result OK

Pressure Fill Real 1006,1

Delta Pressure Real -0,115

Phase Time 0,0

Phase Number 10



Biosigma S.r.l.  
Via Valletta, 6  
00100, Cantalana di Coda (RM), Italy  
Part. IVA 03328440270  
C.F. 02173800281  
Tel. 0661274081 - 06612228