

Brand: Honeywell Fluka
 Product: 30709

Nitric acid min. 65 %, puriss. p.a., Reag. ISO, Reag. Ph. Eur. for determination with dithizone

Parameter	Specification	Units
appearance	complying	
assay	65.0-67.0	%
density (D 20/4)	1.390-1.410	
residue on ignition (SO ₄)	max. 0.0005	%
silver (Ag)	max. 0.000001	%
arsenic (As)	max. 0.000001	%
aluminium (Al)	max. 0.000005	%
barium (Ba)	max. 0.000001	%
beryllium (Be)	max. 0.000001	%
bismuth (Bi)	max. 0.00001	%
calcium (Ca)	max. 0.00001	%
cadmium (Cd)	max. 0.000001	%
cobalt (Co)	max. 0.000001	%
chromium (Cr)	max. 0.000002	%
copper (Cu)	max. 0.000001	%
iron (Fe)	max. 0.00001	%
germanium (Ge)	max. 0.000005	%
mercury (Hg)	max. 0.000001	%
potassium (K)	max. 0.000005	%
lithium (Li)	max. 0.000001	%
magnesium (Mg)	max. 0.00001	%
manganese (Mn)	max. 0.000001	%
molybdenum (Mo)	max. 0.000002	%
sodium (Na)	max. 0.00005	%
nickel (Ni)	max. 0.000002	%
lead (Pb)	max. 0.000001	%

Nitric acid min. 65 %, puriss. p.a., Reag. ISO, Reag. Ph. Eur. for determination with dithizone

Parameter	Specification	Units
strontium (Sr)	max. 0.000001	%
titanium (Ti)	max. 0.00001	%
thallium (Tl)	max. 0.000005	%
vanadium (V)	max. 0.000001	%
zinc (Zn)	max. 0.000005	%
zirconium (Zr)	max. 0.00001	%
heavy metals (as Pb)	max. 0.0002	%
chloride (Cl)	max. 0.00005	%
phosphate (PO ₄)	max. 0.00005	%
sulfate (SO ₄)	max. 0.00005	%
suitab.f.det.w.dithizone	complying	
residual solvents	complying	