

Carlo Erba Reagents S.A.S. Parc d'Activités des Portes Chausée du Vexin - BP 616 27106 Val de Reuil Cedex Tél.:02.32.09.20.00 Fax 02.32.09.20.20 Z.I DE VALDONNE - BP 4 13124 PEYPIN Tél.:04.42.32.41.41 Fax 04.42.72.41.62 Carlo Erba Reagents S.r.l. Via R. Merendi, 22 20010 Cornaredo (Mi) Tel. 02 93 991 90 Fax 02 93 991 001

ISO 9001: 2015

Standard Analysis Certificate

PRODUCT :WATER RPE

CODE :307580 METHOD :6455

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
pH at 25°C	-	5 ÷ 7
Conductivity	μS/cm	<= 0.1
Residue on evaporation	ppm	<= 1
Residue on calcination	ppm	<= 1
Oxidizing substances (O)	mg/l	<= 0.4
Chloride	ppm	<= 0.1
Phosphate	ppm	<= 0.1
Nitrate	ppm	<= 0.1
Sulphate	ppm	<= 0.1
Ag	ppm	<= 0.01
As	ppm	<= 0.01
Au	ppm	<= 0.01
В	ppm	<= 0.01
Ba	ppm	<= 0.01
Be	ppm	<= 0.01
Bi	ppm	<= 0.01
Ca	ppm	<= 0.01
Cd	ppm	<= 0.01
Co	ppm	<= 0.01
Cr	ppm	<= 0.01
Cu	ppm	<= 0.01 <= 0.01
Fe	ppm	<= 0.01 <= 0.01
In	ppm	<= 0.01 <= 0.01
"' K	ppm	<= 0.01 <= 0.01
Li	ppm	<= 0.01 <= 0.01
Mg	ppm	<= 0.01 <= 0.01
Mn		<= 0.01 <= 0.01
Mo	ppm ppm	<= 0.01 <= 0.01
Na		<= 0.01 <= 0.1
Ni	ppm	<= 0.1 <= 0.01
Pb	ppm	<= 0.01 <= 0.01
Si	ppm	<= 0.01 <= 0.01
Sn	ppm	<= 0.01 <= 0.01
Sr	ppm	<= 0.01 <= 0.01
	ppm	<= 0.01 <= 0.01
TI V	ppm	
	ppm	<= 0.01
Zn Z	ppm	<= 0.01
Zr	ppm	<= 0.01
Conductivity during production <= 0.1 µS/cm		
Conform to EN-ISO 3696 grade 3		

Date

:11/01/2020