

# TECHNICAL DATA SHEET ITEM 2491

Mod. AQ 001 Rev. 0 del 25/09/2003

#### **TECHNICAL SPECIFICATIONS**

Trade name	VOLUMETRIC FLASKS WITH SCREW CAP		
Use	No meniscus; calibrated gravimetrically; autoclavable		
Material	PMP TPX®		
Others	Conforms to DIN STD GL 18		

## **DIMENSIONAL SPECIFICATIONS**

Height	mm. 174
External base, diameter	mm. 58
General dimensional tolerance	mm. +/- 0,10
Capacity	ml. 100



#### **GENERAL FEATURES**

CE mark	No
Food contact suitable	Yes

#### **STERILIZATION**

Autoclavability atT +121°C	Yes
Gas (Ethylene Oxide)	Yes
Dry at 160°C	No
Chemical (Formalin)	Yes
Radiations	No
Microwave	Yes

## **CLEANING OPERATIONS**

Slight contamination	Wash with neutral detergent (pH 7)	
Consistent contamination	Wash with alkaline detergent (pH up to 12)	

**Quality Assurance** 

**Technical Direction** 

Massimiliano Capitanio

Guido Borona



# TECHNICAL DATA SHEET ITEM 2491

Mod. AQ 001 Rev. 0 del 25/09/2003

#### **CHEMICAL RESISTANCE**

Categories of substances	Reference	Concen. %	T. 20°	T. 40°	T.° 60
Inorganic acid	Sulphuric acid	98	В	В	В
Organic acid	Benzoic acid	100	A	В	В
Alcohol	Ethanol	100	A	В	В
Aldehydes	Acetaldehyde	100	В	O	-
Inorganic alkali	Sodium hydroxide	50	Α	A	Α
Organic alkali	Aniline	100	В	В	С
Ketone	Acetone	100	В	U	-
Esther	Ethyl acetate	100	A	A	Α
Hydrocarbons, halogenated	Dichloroethylene	100	U	ı	1
Hydrocarbons, aromatic	Benzene	100	В	$\cup$	-
Hydrocarbons, linear	Hexane	100	В	С	-
Supersolvent	Tetrahydrofuran	100	В	В	C

**Legenda** (not applicable if neglected):

- A: fair resistence; exposure (30 days) to the chemical does not cause any damage.
- **B: sufficient resistence**; exposure causes damage of poor importance, which sometimes is only temporary.
- **C: poor resistence**; exposure to chemical is not allowed, or causes immediate permanent damage.

Results of testing with reference materials are to be considered as indication: in case of specific use it is recommended to carry out preliminary testing.

#### **PACKAGING**

Bag	Pcs 20
Box	Pcs 160

**Quality Assurance** 

**Technical Direction** 

Massimiliano Capitanio

Guido Borona

