

Declaration of Conformity

Rev. 1 del 01/09/2006

Noviglio, 08/05/2024

Customer: Société Dominique Dutscher

Item code: 8401300

Description: STERILE CONICAL TEST TUBES PP,50ml, PRINTED GRAD.

Kartell S.P.A., Labware Division, declares under its own responsibility that the above mentioned products, to which this declaration refers, conform to the Company specifications, the inspection plan and the deposited samples.

The articles are pyrogen free.

Limit of ≤ 20EU piece (EU= endotoxin unit).

FDA endotoxin on integral packaging is granted.

Quality Assurance