



DECLARATION OF CONFORMITY

This is to confirm that all our UV instruments and gel documentation systems manufactured by us, including gel documentation systems (Alliance, Essential, Firereader, UVIDoc and UVIsave) and UV based laboratory instruments such as crosslinkers, UV lamps, transilluminators, viewing cabinets, comply with the requirements of the EC Directive 2014/30/EU, 2014/35/EU, 2011/65/EU and EN 61010-1 (electro-magnetic compatibility and low voltage).

The Electro-magnetic susceptibility has been chosen at a level that gains proper operation in residential areas, on business and light industrial premises and on small-scale enterprises, inside as well as outside of the buildings. All places of operation are characterised by their connection to the public low voltage power supply system.

Cambridge, 18 January 2023

For and on behalf of UVIttec Bio Imaging SAS



Ewa Allen,
Sales Office Administrator

UVIttec Cambridge Bio Imaging SAS Registered office: BP42 77601 Marne la Vallee Cedex 3, France
VAT #: FR 88 5621 03838, EORI #: FR 562 103 838 000 65 Mail: uvi@uvitec.co.uk
Web: www.uvitec.co.uk

UK registered address: Bio Imaging Ltd, St John's Innovation Centre, Cowley Road, Cambridge, CB4
0WS

Asia registered Address: Uvitec Bio Imaging, Asia Pte Ltd, 1 North Bridge Road, #22 Floor, Office 09, High Street Centre.
Singapore 179094 Registration UEN 202039819W