AMPLIQON
PCR ENZYMES & REAGENTS

Ampliqon A/S
www.ampliqon.com
info@ampliqon.com

Stenhuggervej 22
5230 Odense M
Denmark

Phone: +45 70201169
VAT No.: DK 26388899

Q-Extract DNA Extraction Hot Start PCR Kit

MADE IN **DENMARK**

Cat. No.	Reactions	Q-Extract DNA Extraction Solution ID: 4900100	TEMPase Hot Start 2x Master Mix A BLUE, 1.5 mM $MgCl_2$ ID: 5200600
A574401	100	1 x 10 ml	1 x 1.25 ml
A574404	500	5 x 10 ml	5 x 1.25 ml

Product description

Q-Extract DNA Extraction Hot Start PCR Kit consists of Q-Extract DNA extraction solution and TEMPase 2x Master Mix A BLUE, which is required for the subsequent PCR.

The Q-Extract DNA Extraction solution is designed for rapid and efficient extraction of PCR-ready DNA from various sample types; mammalian tissues (such as mouse tails and ear snips), plant leaves, saliva, and bacteria. The non-toxic Q-Extract DNA Extraction Solution enables the extraction of DNA from various samples in just 8 minutes. The extraction protocol is divided into two simple heating steps, which is directly followed by PCR using TEMPase Hot Start 2x Master Mix A BLUE. This method is ideal for PCR analysis such as screening and genotyping.

Depending on the sample size, the DNA extraction can be performed in PCR tubes or 1.5 ml tubes, using either a thermocycler or heating block.

TEMPase 2x Master Mix A BLUE is a ready-to-use 2x reaction mix. Each PCR reaction requires 12.5 µl of the master mix. Simply add primers, DNA extract and water to a total reaction volume of 25 µl to successfully carry out PCR.

There is no need to use separate loading dyes. Simply load a portion of the reaction product onto an agarose gel for electrophoresis and subsequent visualization. The blue dye front runs at 400 – 500 bp on a 0.5 – 1.5% agarose gel.

Q-Extract DNA Extraction Hot Start PCR Kit allows for DNA extraction and amplification hereof in less than 2 hours, as compared to ≥1 day with conventional protocols.

Composition of Q-Extract DNA Extraction Solution

- Optimized DNA extraction solution

Composition of TEMPase Hot Start DNA Polymerase 2x Master Mix A BLUE, 1.5 mM $MgCl_2$

- Tris-HCl pH 8.5, $(NH_4)_2S0_4$, 3.0 mM $MgCl_2$, 0.2% Tween® 20
- 0.4 mM of each dNTP
- TEMPase Hot Start DNA Polymerase
- Inert blue dye and stabilizer

Recommended Storage and Stability of Kit Components

Q-Extract DNA Extraction Solution: Long term storage at -20 °C. Product expiry at -20 °C is stated on the label. It can be stored at +4 °C for up to 3 months. Q-Extract DNA Extraction Solution tolerates up to 20 freeze-thaw cycles. It is recommended to aliquot the Q-Extract into smaller volumes.

TEMPase 2x Master Mix A BLUE: Long term storage at -20 °C. Product expiry at -20 °C is stated on the label. It can be stored at +4 °C for up to 6 months.

Quality Control

Each batch of Q-Extract DNA Extraction Solution is functionally tested.
TEMPase Hot Start DNA Polymerase is functionally tested for contaminating activities, with no traces of endonuclease activity, nicking activity or exonuclease activity.

Extraction Protocol

Preparation of DNA extraction should be performed in a separate area from that used for setting up the PCR.

1. Thaw Q-Extract DNA Extraction Solution. For the first time use, aliquot the Q-Extract DNA Extraction solution into smaller volumes. (Q-Extract DNA Extraction Solution has a cloudy appearance).
2. Add your sample to a tube containing 100 µl Q-Extract DNA Extraction Solution. Recommended sample sizes are shown in table 1.
3. Vortex the tube containing the sample and the DNA extraction solution for 15 sec.
4. Transfer the tube to a heat block or a thermal cycler and incubate for

    1. 65 °C for 6 min
    2. 98 °C for 2 min
    3. 4 °C (or cool down on ice)

The DNA extract is now ready for PCR. See PCR protocol in table 2.

DNA extracts are stable at -20 °C for one week or long term at -80 °C.

Table 1. Sample sizes

Sample	Q-Extract DNA Extraction Solution	
	100 µl	500 µl
Tissue*	0.5 – 10 mg	10 – 50 mg
Plant**	2 – 10 mg	10 – 50 mg
E. coli	1 colony (Φ 0.5 - 2 mm)	1 colony (Φ 0.5 - 5 mm)
Saliva	10 – 20 µl	50 - 100 µl

* Examples of tested tissues include mouse tail snip, mouse organs and chicken breast.
**Examples of tested plant materials include leaves from stinging nettle and ivy.

PCR Protocol

This protocol serves as a guideline to ensure optimal PCR results when using TEMPase 2x Master Mix A BLUE. Optimal reaction conditions such as incubation times, temperatures, and amount of template DNA may vary and must be determined individually.

1. Thaw TEMPase 2x Master Mix A BLUE and primers.
   It is important to thaw the solutions completely and mix thoroughly before use to avoid localized concentrations of salts.
   Keep all components on ice.
2. Prepare a reaction mix. Table 2 shows the reaction set up for a final volume of 25 µL. If desired, the reaction size may be scaled up or down.

DOMINIQUE DUTSCHER SAS

Table 2. Reaction components (reaction mix and template DNA)

Component	Vol./reaction*	Final concentration*
2x Master Mix	12.5 µl	1x
25 mM $MgCl_2$	0 µl (0 – 3 µl)	1.5 mM (1.5 – 4.5 mM)
Primer A (10 µM)	0.5 µl (0.25 – 2.5 µl)	0.2 µM (0.1 – 1.0 µM)
Primer B (10 µM)	0.5 µl (0.25 – 2.5 µl)	0.2 µM (0.1 – 1.0 µM)
PCR-grade H_2O	X µl	-
DNA Extract**	2 - 5 µl	Variable
TOTAL volume	25 µl	-

* Suggested starting conditions; theoretically used conditions in brackets
** If the PCR yields are poor or one experience no bands, it might help to dilute the DNA extract 1:10. DNA extracts from plant leaves should be diluted 1:10 or 1:100, especially when analysing chloroplast DNA.

3. Mix gently.
4. Add extracted DNA to the individual tubes containing the reaction mix.
5. Program the thermal cycler according to the manufacturer´s instructions. Example in table 3.
   For maximum yield and specificity, temperatures and cycling times should be optimized for each new template target or primer pair.
6. Place the tubes in the thermal cycler and start the reaction.
7. At the end of the run, simply load a portion of the reaction product (e.g. 10 µl) onto an agarose gel for analysis.

Table 3. Three-step PCR program

Cycles	Duration of cycle	Temperature
1	15 minutes	95 °C
25 - 35	20 – 30 seconds 20 – 40 seconds 30 seconds	95 °C 50 – 65 °C 72 °C
1	5 minutes	72 °C

Notes:

- For genotyping of fish fins and other applications please visit our website.

- The final $MgCl_2$ concentration of TEMPase 2x Master Mix A BLUE is 1.5 mM. In some PCR applications, more than 1.5 mM $MgCl_2$ is required for best results. Use 25 mM to adjust the Mg^{2+} concentration according to table 4.

Table 4. Additional volume (µl) of $MgCl_2$ per 25 µl reaction:

Final $MgCl_2$ conc. in reaction (mM)	1.5	2.0	2.5	3.0	3.5	4.0	4.5
Volume of 25 mM $MgCl_2$	0	0.5	1	1.5	2	2.5	3

For Research Use Only. Not for use in diagnostics procedures.

Other product sizes, combinations and customized solutions are available. Please look at www.ampliqon.com or ask for our complete product list for PCR Enzymes. For customized solutions please contact us.

Made in Denmark

Issued 08/2025

DOMINIQUE DUTSCHER SAS