

# Standard Analysis Certificate

**PRODUCT** :BUFFER SOLUTION pH 10.06 BLUE RPE**CODE** :486810

TEST	U.M.	SPECIFICATION
Appearance	-	Blue clear solution
pH	unité pH	10.01 - 10.11
Temperature of measurement	°C	19 - 21

Accuracy : +/- 0.05 pH units - Standardized against NIST reference materials.

Date

:11/01/2020

**QUALITY CONTROL RESPONSIBLE****B. COULANGE (VDR)**