

Standard Analysis Certificate

PRODUCT :HYDROCHLORIC ACID 0.01 N (0.01 mol/L) RPE**CODE** :404260

TEST	U.M.	SPECIFICATION
Description Assay (potentiometry)	- N	Clear colourless liquid 0.00998 - 0.01002

This volumetric solution is checked by sodium hydroxyde solution, tested with potassium hydrogen phthalate volumetric standard traceable to the SRM 84L NIST (National Institute of Standards and Technology)

DOMINIQUE DUTSCHER SAS

Date

:11/01/2020

QUALITY CONTROL RESPONSIBLE**B. COULANGE (VDR)**