

Standard Analysis Certificate

PRODUCT :WATER PLUS RS For HPLC
CODE :412140
METHOD :7684

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity during production	µS/cm	<= 0.1
Residue on evaporation	ppm	<=0.5
Total organic carbon	ppm	<=0.1
CO2	-	Not detectable
Nitrate	ppm	<= 0.1
Heavy metals (Pb)	ppm	<= 0.1
Fluorescence	-	-
at 254 nm	ppb	<=1
at 365 nm	ppb	<=0.5
UV Abs.max elut.peak	-	-
at 210 nm	mAu	<=5
at 220 nm	mAu	<=3
at 254 nm	mAu	<=0.5

Filtered at 0.1 micron

DOMINIQUE DUTSCHER SAS

Date :11/01/2020

QUALITY CONTROL RESPONSIBLE
B. COULANGE (VDR)