

Standard Analysis Certificate

PRODUCT :BUFFER SOLUTION pH 3 RPE**CODE** :486250

| TEST | U.M. | SPECIFICATION |
|----------------------------|----------|---------------|
| Clear,colourless solution | - | Conform |
| pH | unité pH | 2.98 - 3.02 |
| Temperature of measurement | °C | 19 - 21 |

Accuracy : +/- 0.02 pH units - Standardized against NIST reference materials.

Date

:11/01/2020

QUALITY CONTROL RESPONSIBLE**B. COULANGE (VDR)**