

# Standard Analysis Certificate

**PRODUCT** :BUFFER SOLUTION pH 6 RPE**CODE** :486330

TEST	U.M.	SPECIFICATION
Clear,colourless solution	-	Conform
pH	unité pH	5.98 - 6.02
Temperature of measurement	°C	19 - 21

Accuracy : +/- 0.02 pH units - Standardized against NIST reference materials.

Date

:11/01/2020

**QUALITY CONTROL RESPONSIBLE****B. COULANGE (VDR)**