

Standard Analysis Certificate

PRODUCT :BUFFER SOLUTION pH 7 Yellow NORMEX RPE
CODE :486431
METHOD :9922

TEST	U.M.	SPECIFICATION
Description	-	Yellow clear liquid
Identification	-	Positive
pH at 20° C	-	6.95 ÷ 7.05

Guaranteed pH range determined referring to Standard NIST.

DOMINIQUE DUTSCHER SAS

Date

:11/01/2020

QUALITY CONTROL RESPONSIBLE

B. COULANGE (VDR)