

# Standard Analysis Certificate

**PRODUCT** :BUFFER SOLUTION pH 10 +/- 0.05**CODE** :486610

TEST	U.M.	SPECIFICATION
Clear,colourless solution	-	Conform
pH	unité pH	9.95 - 10.05
Temperature of measurement	°C	19 - 21

Accuracy : +/- 0.05 pH units - Standardized against NIST reference materials.

DOMINIQUE DUTSCHER SAS

Date

:11/01/2020

**QUALITY CONTROL RESPONSIBLE****B. COULANGE (VDR)**