

# Standard Analysis Certificate

**PRODUCT** :BUFFER SOLUTION pH 4 RED RPE**CODE** :486760

TEST	U.M.	SPECIFICATION
Appearance	-	Red clear solution
pH	unité pH	3.98 - 4.02
Temperature of measurement	°C	19 - 21

Accuracy : +/- 0.02 pH units - Standardized against NIST reference materials.

Date

:11/01/2020

**QUALITY CONTROL RESPONSIBLE****B. COULANGE (VDR)**