

Standard Analysis Certificate

PRODUCT :BUFFER SOLUTION pH 7 GREEN RPE**CODE** :486790

TEST	U.M.	SPECIFICATION
Appearance	-	Green clear solution
pH	unité pH	6.98 - 7.02
Temperature of measurement	°C	19 - 21

Accuracy : +/- 0.02 pH units - Standardized against NIST reference materials.

Date

:11/01/2020

QUALITY CONTROL RESPONSIBLE**B. COULANGE (VDR)**