## **EU Declaration of Conformity**

Product Laboratory Equipment File Number TFR924051 007

**Manufacturer** Cole-Parmer Ltd

Beacon Road

Stone, Staffordshire

ST15 0SA

United Kingdom

## This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration Dri-Block Heater with RS232 output

BH-250D series

The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation:

Low Voltage Directive2014/35/EUEMC Directive2014/30/EURoHS Directive2011/65/EC

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

IEC/EN 61010-1:2010	Safety requirements for electrical equipment for measurement, control and laboratory use. Part 1: General requirements.
IEC/EN 61010-2-010:2014	Particular requirements for laboratory equipment for the heating of materials.
IEC/EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use. EMC requirements. Part 1: General requirements (Class A).

## Signed for and on behalf of the above manufacturer

**Additional Information** Year of CE Marking: 1996

**Place of Issue** Stone, Staffordshire, UK

Date of Issue03 January 2017Authorised RepresentativeSteve Marriott

**Title** Technical Director

Signature

Catalogue Number	Model	Description
36620-16	BH-250D-3-HT-RS	Digital Heating Block, triple insert, 200°C 230 V RS-232 output
99965-78	BH-250D-2-HT-RS	Digital Heating Block, dual insert, 200°C 230 V RS-232 output