



Declaration of conformity for the NanoPhotometer® (C40/N50/N60/NP80/N120)

This is to certify that the Implen NanoPhotometer® conforms to the requirements of the following directives:

2014/35/EU	Low Voltage Equipment Safety Directive
2014/30/EU	Electromagnetic compatibility (EMC) directive
IEC 60529	Protection class IP20
2011/65/EU	Restrictions on the use of certain Hazardous Substances in Electrical and Electronic Equipment (ROHS)
2012/19/EU	EC Directive on Waste Electrical and Electronic Equipment (WEEE) 2003/108/EC & 2008/34/EC. By ensuring this product is disposed of correctly, you will help prevent potential negative consequences for the environment and human health, which could otherwise be caused by inappropriate waste handling of this product.

FCC 47 CFR Part15 §15.107 and §15.109

EN 301 489-1 V1.9.2 Radio and ancillary equipment for portable use (portable equipment); EUT Operating frequency range: 2.4 – 2.4835 GHz

EN 300 328 V1.8.1 Electromagnetic compatibility and Radio spectrum Matters (ERM); Wideband transmission systems; Data transmission equipment operating in the 2,4 GHz ISM band and using wide band modulation techniques; Harmonized EN covering the essential requirements of article 3.2 of the R&TTE Directive

EN 301 489-17 V2.2.1 Electromagnetic compatibility and Radio Spectrum Matters (ERM)

IEC 62133 and UN38.3 Battery certification and transport test

Standards to which conformity is declared, where relevant, are as follows:

IEC/EN 61010-1:2012 Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements.

EN61326-1:2013 Electromagnetic compatibility- generic emission standard electrical equipment for measurement, control, and laboratory use.

Signed:

A handwritten signature in blue ink, appearing to read "Th. Sahiri".

Dr. Thomas Sahiri
Managing Director
Implen GmbH