CERTIFICATE OF COMPLIANCE

Certificate No. : FQC2516501

Certificate Holder and Address	:	**HANGZHOU ALLSHENG INSTRUMENTS CO., LTD.** *BUILDING 9 NO.7 OF ZHUANTANG SCIENCE AND TECHNOLOGY ECONOMIC ZONE, XIHU DISTRICT, HANGZHOU CITY, ZHEJIANG 310024, CHINA*
Name and Address of Manufacturer	:	**HANGZHOU ALLSHENG INSTRUMENTS CO., LTD.** *BUILDING 9 NO.7 OF ZHUANTANG SCIENCE AND TECHNOLOGY ECONOMIC ZONE, XIHU DISTRICT, HANGZHOU CITY, ZHEJIANG 310024, CHINA*
Product Name	:	Microplate Centrifuge
Product Model	:	MPC-P25, MPC-P25A, C2001, C2002-FIS
Applied Directive/s	:	*2014/30/EU Electromagnetic Compatibility* *2014/35/EU Low Voltage Directive*
Applied Standard/s	:	EN 61010-1:2010+A1:2019, EN 61010-2-020:2020, EN 61010-2-101:2022+A11:2022, EN 61326-1:2021, EN 61326-2-6:2021
Base of Attestation	:	Declaration of Conformity *ALLSHENG-2025A09*

FQC STANDARD certifies with the presented technical documentation that the type presented above for conformity assessment complies with 2014/30/EU Electromagnetic Compatibility Regulation and 2014/35/EU Low Voltage Directive. This certificate of conformity does not relieve the manufacturer of the obligation to issue a declaration of conformity and comply with other relevant regulations. This certificate of conformity is issued at the request of the manufacturer.

Certificate Issue Date : **18.04.2025**

Certificate Expiry Date : **18.04.2030**

FQC STANDARD UYGUNLUK DEĞERLENDİRME A.Ş. FQC STANDARD PRODUCT

General Manager