Attestation of Conformity

ATTESTATION OF CONFORMITY

Name and Address of Attestation Holder:
HANGZHOU ALLSHENG INSTRUMENTS CO., LTD.
BUILDING 9 NO.7 OF ZHUANTANG SCIENCE AND TECHNOLOGY ECONOMIC ZONE, XIHU DISTRICT, HANGZHOU CITY, ZHEJIANG 310024, CHINA.

Name and Address of Manufacturer:
HANGZHOU ALLSHENG INSTRUMENTS CO., LTD.
BUILDING 9 NO.7 OF ZHUANTANG SCIENCE AND TECHNOLOGY ECONOMIC ZONE, XIHU DISTRICT, HANGZHOU CITY, ZHEJIANG 310024, CHINA.

Brand:
N/A

Product Name:
Micro-Spectrophotometer

Product Model:
Nano-400A

Document Number
CE-1437-08-151123

Date of Issue:
15.11.2023

Expiration Date:
15.11.2028

Test Report Number:
ALLSHENG-2023A08

Test Required:
BS EN IEC 61010-1:2010 +A1:2019
BS EN IEC 61010-2-101:2017
BS EN IEC 61326-1:2021
BS EN IEC 61326-2-6:2021

The product meets the technical requirement of the above standards as mentioned in the reference test reports and hence fulfils the technical requirements of the following directives

2014/35/EU Low Voltage Directive
2014/30/EU Electromagnetic Compatibility Directive

This document is only valid for the equipment and configuration described, in conjunction with the test data detailed above reference test reports. Document was issued on voluntary basis and does not imply meeting Notified Body conformity assessment procedure for the product.
The CE Mark, under the responsibility of the manufacturer or the importer, after completion of an EC Declaration of Conformity and compliance with all relevant EC Directives.

SIGNATURE

CGS TEST HİZMETLERİ TEKNİK KONTROL VE BELGELENDİRME ANONİM ŞİRKETİ
Ferhatpaşa Mahallesi 23. Sk. No:17/1 Ataşehir İstanbul/TÜRKİYE
www.cgstestmerkezi.com