

Standard Analysis Certificate

PRODUCT :CONDUCTIVITY STANDARD 10 µS/cm RS
CODE :575010
METHOD :10448

TEST	U.M.	SPECIFICATION
Description	-	Clear colourless liquid
Identification	-	Positive
Conductivity at 25°C	µS/cm	9.80 ÷ 10.20

Traceable to Standard NIST N° 999a
 Confidence interval: ± 0.02 µS/cm

DOMINIQUE DUTSCHER SAS

Date :11/01/2020

QUALITY CONTROL RESPONSIBLE
B. COULANGE (VDR)